

Global Superconducting MRI Equipment Market Growth 2026-2032

<https://marketpublishers.com/r/S3BCC6C3FE4BEN.html>

Date: October 2026

Pages: 118

Price: US\$ 3,660.00 (Single User License)

ID: S3BCC6C3FE4BEN

Abstracts

The global Superconducting MRI Equipment market size is predicted to grow from US\$ 6383 million in 2025 to US\$ 9236 million in 2032; it is expected to grow at a CAGR of 5.5% from 2026 to 2032.

Superconducting MRI Equipment refers to magnetic resonance imaging systems that use superconducting magnet technology to generate a stable, high-intensity static magnetic field for non-ionizing diagnostic imaging. The system architecture typically integrates a superconducting magnet, cryogenic and cooling system, gradient subsystem, radio-frequency transmit and receive components, dedicated coils, patient handling system, image reconstruction hardware, clinical software, and workflow applications. Commercial systems span sub-1.5T, 1.5T, 3.0T and ultra-high-field configurations above 3.0T, with 1.5T and 3.0T forming the principal clinical platforms. The product scope covers conventional liquid-helium and sealed low-helium magnet architectures, closed-bore and open-bore designs, and fixed or mobile system configurations. Superconducting MRI Equipment is used for neurological, musculoskeletal, cardiovascular, abdominal, pelvic, breast, oncological and other diagnostic examinations, while high-performance systems also support functional imaging, quantitative imaging, advanced diffusion imaging and clinical research.

Key Findings

1.5T and 3T systems remain the core clinical field strengths

Low helium sealed magnets are reshaping MRI lifecycle economics

AI reconstruction increasingly shifts competition toward software enabled productivity

Ultra high field systems extend MRI into new clinical frontiers

Regional MRI access remains highly uneven across healthcare systems

Market Trends

Superconducting MRI Equipment is moving from competition centered primarily on magnet field strength toward integrated competition in magnet architecture, gradient performance, radio-frequency technology, reconstruction algorithms, workflow automation and lifecycle efficiency. Deep-learning reconstruction is increasingly embedded at the system platform level, allowing manufacturers to improve signal-to-noise performance, spatial resolution or scanning speed without relying solely on higher magnetic field strength. At the same time, automated positioning, protocol selection, coil management and examination workflow are reducing operator complexity and supporting higher patient throughput. Magnet design is undergoing another structural shift: sealed low-helium systems, often positioned as helium-independent or helium-free in routine operation, reduce dependence on conventional helium infrastructure and can simplify site preparation. Wider bores, flexible coils, lower acoustic burden and patient-centered workflow are becoming important purchasing considerations alongside image quality. At the technological frontier, whole-body 5T and clinical 7T platforms are expanding the development space for ultra-high-field imaging, although their near-term role remains more concentrated in advanced clinical and research environments than routine imaging.

Market Dynamics

Drivers

Demand for Superconducting MRI Equipment is supported by the continuing importance of MRI in neurological, musculoskeletal, oncological, cardiovascular and abdominal imaging, together with expanding requirements for higher-resolution and quantitative diagnosis. Unlike imaging modalities based on ionizing radiation, MRI provides strong soft-tissue contrast while supporting a broad range of anatomical and functional imaging techniques. Healthcare providers are also under pressure to increase scanner utilization and manage growing imaging workloads, making shorter examinations, automated workflows and higher system uptime increasingly valuable. Improvements in deep-learning reconstruction, compressed sensing, gradient performance and coil technology are enabling more examinations to be completed within practical clinical timeframes. In

parallel, aging installed bases in mature healthcare systems create recurring replacement and upgrade demand, while lower equipment availability in many emerging healthcare markets continues to support long-term demand for new capacity.

Restraints

The principal constraints remain high capital expenditure, demanding installation conditions and the lifecycle costs associated with operating sophisticated imaging infrastructure. Conventional superconducting systems may require significant site preparation involving magnetic shielding, cooling, power infrastructure and cryogenic safety engineering, while system downtime can materially affect imaging department economics. High-performance 3.0T and ultra-high-field systems also require stronger technical capabilities in magnet homogeneity, gradient design, radio-frequency management, reconstruction and clinical protocol development. Purchasing decisions therefore depend not only on acquisition price but also on installation costs, service capability, expected examination volumes and long-term upgradeability. The increasing sophistication of software and hardware additionally raises the importance of qualified radiographers, engineers and clinical application specialists, which can limit adoption where trained personnel and service infrastructure remain insufficient.

Opportunities

The largest structural opportunities are developing around replacement of legacy equipment, expansion of MRI access, low-helium magnet architecture and software-driven productivity upgrades. Sealed magnet technology can reduce dependence on conventional helium handling infrastructure and make installation more practical in locations where conventional siting is difficult. Compact footprints and mobile configurations can further extend MRI capacity beyond large tertiary hospitals. At the same time, artificial intelligence creates opportunities to improve existing 1.5T and 3.0T platforms through reconstruction, positioning, protocol automation and workflow optimization rather than relying exclusively on higher-field hardware. Research-oriented 5T and 7T systems create another technology opportunity in high-resolution neurological, musculoskeletal, metabolic and multinuclear imaging. Over the longer term, successful translation of these advanced capabilities into standardized clinical protocols could broaden the commercial addressability of ultra-high-field platforms.

Challenges

The market faces an increasingly complex balance between technological performance and clinical economic value. Higher field strength can provide greater signal potential,

but ultra-high-field platforms introduce additional challenges in radio-frequency uniformity, specific absorption rate management, gradient performance, coil design and protocol optimization. Artificial-intelligence-based reconstruction and automation must also maintain stable performance across anatomies, patient populations, acquisition conditions and clinical workflows. At the commercial level, manufacturers must simultaneously manage hardware innovation, regulatory compliance, cybersecurity, software upgrades, field service and long equipment lifecycles. Buyers are becoming more focused on total cost of ownership and measurable workflow improvement, which raises the threshold for premium system differentiation. Competition therefore increasingly depends on whether new technology can be translated into reproducible clinical productivity and diagnostic value rather than technical specifications alone.

Industry Chain Analysis

The upstream industry chain for Superconducting MRI Equipment is technology intensive and includes superconducting materials and magnet components, cryogenic and thermal-management systems, gradient coils and amplifiers, radio-frequency components, high-density receiving coils, power electronics, precision mechanical structures, sensors, computing hardware and specialized electronic components. Magnet design determines the fundamental field environment, while gradient and radio-frequency systems strongly influence acquisition speed, spatial resolution and advanced sequence capability. As sealed low-helium architectures gain importance, cryogenic engineering, thermal insulation and magnet reliability are becoming increasingly differentiated upstream capabilities. High-performance computing components and artificial-intelligence acceleration are also gaining value as reconstruction moves from conventional algorithms toward computationally intensive deep-learning techniques.

The midstream combines system engineering, magnet and gradient integration, radio-frequency architecture, image reconstruction, clinical sequence development, workflow software, regulatory validation and final manufacturing. Value creation has therefore shifted from hardware integration alone toward a combination of core hardware, proprietary imaging technologies, software ecosystems and lifecycle service. Downstream value is realized through hospitals, diagnostic imaging centers, specialty clinics and research institutions, where installation engineering, application training, preventive maintenance, software upgrades and system availability materially influence purchasing decisions. This structure favors manufacturers capable of maintaining long-term product platforms and service networks because MRI systems have long operating lives and can receive substantial software, coil and workflow upgrades during their

service period.

Segment Insights

By magnetic field strength, 1.5T and 3.0T systems represent the commercial center of the Superconducting MRI Equipment portfolio. The 1.5T category addresses broad routine diagnostic requirements and is particularly relevant where hospitals prioritize clinical versatility, installation flexibility and operating economics. The 3.0T category increasingly serves advanced neurological, musculoskeletal, cardiovascular and research-oriented applications where higher signal performance supports finer spatial resolution or more demanding acquisition techniques. Systems below 1.5T provide differentiated opportunities in selected accessibility, installation and patient-comfort scenarios, while systems above 3.0T remain a more specialized high-performance segment. The introduction of whole-body 5T and clinical 7T systems demonstrates that ultra-high-field MRI is gradually extending beyond purely experimental platforms.

From the perspective of magnet cooling architecture, conventional liquid-helium systems continue to represent an important installed technology base, while sealed low-helium designs are becoming a strategically important direction for new platforms. Their value proposition extends beyond helium consumption to installation engineering, quench infrastructure, uptime risk and lifecycle management. Closed-bore systems remain the principal architecture across mainstream 1.5T and 3.0T product portfolios, while open-bore configurations address selected patient-comfort and positioning requirements. Fixed systems form the core deployment model, while mobile superconducting MRI offers a complementary solution for temporary capacity, regional service expansion and flexible diagnostic deployment.

Downstream Market Opportunities

Hospitals remain the anchor application environment because they require broad anatomical coverage, emergency and inpatient support, complex examinations and integration with multidisciplinary clinical pathways. Diagnostic imaging centers place greater emphasis on examination throughput, predictable operating costs, standardized protocols and patient experience, making workflow automation and rapid reconstruction particularly relevant. Specialty clinics can benefit from application-focused configurations, wide-bore designs and streamlined examination protocols, while academic medical centers and clinical research institutions form the principal demand base for high-gradient, high-field and ultra-high-field platforms. Mobile systems and more installation-friendly magnet architectures provide an additional route for expanding

diagnostic capacity where permanent infrastructure or patient volumes do not initially justify a conventional fixed installation.

Regional Insights

North America, Europe and the more developed healthcare systems of Asia-Pacific represent mature Superconducting MRI Equipment markets characterized by established installed bases, frequent technology upgrades and demand for advanced 3.0T and research-oriented systems. Replacement purchasing in these markets increasingly emphasizes examination productivity, artificial intelligence, patient comfort, lower energy and helium dependence, and the ability to extend system life through hardware and software upgrades. At the same time, the Asia-Pacific region presents particularly wide differences in equipment accessibility. OECD data indicate approximately 57 MRI units per million population in Japan, while several developing Asia-Pacific markets report fewer than one unit per million population. These figures cover MRI equipment overall rather than superconducting systems specifically, but they highlight the substantial disparity in diagnostic imaging infrastructure across the region.

This disparity produces two distinct regional opportunities. Highly equipped markets are driven primarily by replacement, premium system upgrades and specialized high-field applications, whereas lower-density markets have greater potential for capacity expansion and first-time installations. China represents an increasingly important technology and manufacturing base, with domestic suppliers expanding from mainstream 1.5T platforms toward 3.0T, helium-independent and ultra-high-field systems. Other emerging markets in Asia, the Middle East, Latin America and parts of Africa are more sensitive to acquisition cost, installation complexity, service coverage and operating reliability, creating opportunities for compact, lower-helium and workflow-efficient systems rather than competition based solely on maximum field strength.

Report Scope

LP Information, Inc. (LPI) ' newest research report, the "Superconducting MRI Equipment Industry Forecast" looks at past sales and reviews total world Superconducting MRI Equipment sales in 2025, providing a comprehensive analysis by region and market sector of projected Superconducting MRI Equipment sales for 2026 through 2032. With Superconducting MRI Equipment sales broken down by region, market sector and sub-sector, this report provides a detailed analysis in US\$ millions of the world Superconducting MRI Equipment industry.

This Insight Report provides a comprehensive analysis of the global Superconducting MRI Equipment landscape and highlights key trends related to product segmentation, company formation, revenue, and market share, latest development, and M&A activity. This report also analyzes the strategies of leading global companies with a focus on Superconducting MRI Equipment portfolios and capabilities, market entry strategies, market positions, and geographic footprints, to better understand these firms' unique position in an accelerating global Superconducting MRI Equipment market.

This Insight Report evaluates the key market trends, drivers, and affecting factors shaping the global outlook for Superconducting MRI Equipment and breaks down the forecast by Type, by Application, geography, and market size to highlight emerging pockets of opportunity. With a transparent methodology based on hundreds of bottom-up qualitative and quantitative market inputs, this study forecast offers a highly nuanced view of the current state and future trajectory in the global Superconducting MRI Equipment.

This report presents a comprehensive overview, market shares, and growth opportunities of Superconducting MRI Equipment market by product type, application, key manufacturers and key regions and countries.

Market Segmentation

Segmentation by Type:

1.5T

3.0T

Others

Segmentation by Pore ??Structure:

Closed-Bore Superconducting MRI Systems

Open-Bore Superconducting MRI Systems

Segmentation by Installation and Mobility:

Fixed Superconducting MRI Systems

Mobile Superconducting MRI Systems

Segmentation by Application:

Hospitals

Imaging Centers

Other

This report also splits the market by region:

Americas

United States

Canada

Mexico

Brazil

APAC

China

Japan

Korea

Southeast Asia

India

Australia

Europe

Germany

France

UK

Italy

Russia

Middle East & Africa

Egypt

South Africa

Israel

Turkey

GCC Countries

The below companies that are profiled have been selected based on inputs gathered from primary experts and analysing the company's coverage, product portfolio, its market penetration.

Siemens

Philips

GE HealthCare

Canon Medical

FUJIFILM

United Imaging

Neusoft Medical

XGY

Anke

Alltech

Wandong

BASDA

Esaote S.p.A.

Key Questions Addressed in this Report

What is the 10-year outlook for the global Superconducting MRI Equipment market?

What factors are driving Superconducting MRI Equipment market growth, globally and by region?

Which technologies are poised for the fastest growth by market and region?

How do Superconducting MRI Equipment market opportunities vary by end market size?

How does Superconducting MRI Equipment break out by Type, by Application?

Contents

1 SCOPE OF THE REPORT

- 1.1 Market Introduction
- 1.2 Years Considered
- 1.3 Research Objectives
- 1.4 Market Research Methodology
- 1.5 Research Process and Data Source
- 1.6 Economic Indicators
- 1.7 Currency Considered
- 1.8 Market Estimation Caveats

2 EXECUTIVE SUMMARY

2.1 World Market Overview

- 2.1.1 Global Superconducting MRI Equipment Annual Sales 2021-2032
- 2.1.2 World Current & Future Analysis for Superconducting MRI Equipment by Geographic Region, 2021, 2025 & 2032
- 2.1.3 World Current & Future Analysis for Superconducting MRI Equipment by Country/Region, 2021, 2025 & 2032
- 2.2 Superconducting MRI Equipment Segment by Type
 - 2.2.1 1.5T
 - 2.2.2 3.0T
 - 2.2.3 Others
 - 2.2.4 Superconducting MRI Equipment Sales by Type
 - 2.2.4.1 Global Superconducting MRI Equipment Sales Market Share by Type (2021-2026)
 - 2.2.4.2 Global Superconducting MRI Equipment Revenue and Market Share by Type (2021-2026)
 - 2.2.4.3 Global Superconducting MRI Equipment Sale Price by Type (2021-2026)
- 2.3 Superconducting MRI Equipment Segment by Pore ??Structure
 - 2.3.1 Closed-Bore Superconducting MRI Systems
 - 2.3.2 Open-Bore Superconducting MRI Systems
 - 2.3.3 Superconducting MRI Equipment Sales by Pore ??Structure
 - 2.3.3.1 Global Superconducting MRI Equipment Sales Market Share by Pore ??Structure (2021-2026)
 - 2.3.3.2 Global Superconducting MRI Equipment Revenue and Market Share by Pore ??Structure (2021-2026)

2.3.3.3 Global Superconducting MRI Equipment Sale Price by Pore ??Structure (2021-2026)

2.4 Superconducting MRI Equipment Segment by Installation and Mobility

2.4.1 Fixed Superconducting MRI Systems

2.4.2 Mobile Superconducting MRI Systems

2.4.3 Superconducting MRI Equipment Sales by Installation and Mobility

2.4.3.1 Global Superconducting MRI Equipment Sales Market Share by Installation and Mobility (2021-2026)

2.4.3.2 Global Superconducting MRI Equipment Revenue and Market Share by Installation and Mobility (2021-2026)

2.4.3.3 Global Superconducting MRI Equipment Sale Price by Installation and Mobility (2021-2026)

2.5 Superconducting MRI Equipment Segment by Application

2.5.1 Hospitals

2.5.2 Imaging Centers

2.5.3 Other

2.5.4 Superconducting MRI Equipment Sales by Application

2.5.4.1 Global Superconducting MRI Equipment Sale Market Share by Application (2021-2026)

2.5.4.2 Global Superconducting MRI Equipment Revenue and Market Share by Application (2021-2026)

2.5.4.3 Global Superconducting MRI Equipment Sale Price by Application (2021-2026)

3 GLOBAL BY COMPANY

3.1 Global Superconducting MRI Equipment Breakdown Data by Company

3.1.1 Global Superconducting MRI Equipment Annual Sales by Company (2021-2026)

3.1.2 Global Superconducting MRI Equipment Sales Market Share by Company (2021-2026)

3.2 Global Superconducting MRI Equipment Annual Revenue by Company (2021-2026)

3.2.1 Global Superconducting MRI Equipment Revenue by Company (2021-2026)

3.2.2 Global Superconducting MRI Equipment Revenue Market Share by Company (2021-2026)

3.3 Global Superconducting MRI Equipment Sale Price by Company

3.4 Key Manufacturers Superconducting MRI Equipment Producing Area Distribution, Sales Area, Product Type

3.4.1 Key Manufacturers Superconducting MRI Equipment Product Location Distribution

- 3.4.2 Players Superconducting MRI Equipment Products Offered
- 3.5 Market Concentration Rate Analysis
 - 3.5.1 Competition Landscape Analysis
 - 3.5.2 Concentration Ratio (CR3, CR5 and CR10) & (2024-2026)
- 3.6 New Products and Potential Entrants
- 3.7 Market M&A Activity & Strategy

4 WORLD HISTORIC REVIEW FOR SUPERCONDUCTING MRI EQUIPMENT BY GEOGRAPHIC REGION

- 4.1 World Historic Superconducting MRI Equipment Market Size by Geographic Region (2021-2026)
 - 4.1.1 Global Superconducting MRI Equipment Annual Sales by Geographic Region (2021-2026)
 - 4.1.2 Global Superconducting MRI Equipment Annual Revenue by Geographic Region (2021-2026)
- 4.2 World Historic Superconducting MRI Equipment Market Size by Country/Region (2021-2026)
 - 4.2.1 Global Superconducting MRI Equipment Annual Sales by Country/Region (2021-2026)
 - 4.2.2 Global Superconducting MRI Equipment Annual Revenue by Country/Region (2021-2026)
- 4.3 Americas Superconducting MRI Equipment Sales Growth
- 4.4 APAC Superconducting MRI Equipment Sales Growth
- 4.5 Europe Superconducting MRI Equipment Sales Growth
- 4.6 Middle East & Africa Superconducting MRI Equipment Sales Growth

5 AMERICAS

- 5.1 Americas Superconducting MRI Equipment Sales by Country
 - 5.1.1 Americas Superconducting MRI Equipment Sales by Country (2021-2026)
 - 5.1.2 Americas Superconducting MRI Equipment Revenue by Country (2021-2026)
- 5.2 Americas Superconducting MRI Equipment Sales by Type (2021-2026)
- 5.3 Americas Superconducting MRI Equipment Sales by Application (2021-2026)
- 5.4 United States
- 5.5 Canada
- 5.6 Mexico
- 5.7 Brazil

6 APAC

6.1 APAC Superconducting MRI Equipment Sales by Region

6.1.1 APAC Superconducting MRI Equipment Sales by Region (2021-2026)

6.1.2 APAC Superconducting MRI Equipment Revenue by Region (2021-2026)

6.2 APAC Superconducting MRI Equipment Sales by Type (2021-2026)

6.3 APAC Superconducting MRI Equipment Sales by Application (2021-2026)

6.4 China

6.5 Japan

6.6 South Korea

6.7 Southeast Asia

6.8 India

6.9 Australia

6.10 China Taiwan

7 EUROPE

7.1 Europe Superconducting MRI Equipment by Country

7.1.1 Europe Superconducting MRI Equipment Sales by Country (2021-2026)

7.1.2 Europe Superconducting MRI Equipment Revenue by Country (2021-2026)

7.2 Europe Superconducting MRI Equipment Sales by Type (2021-2026)

7.3 Europe Superconducting MRI Equipment Sales by Application (2021-2026)

7.4 Germany

7.5 France

7.6 UK

7.7 Italy

7.8 Russia

8 MIDDLE EAST & AFRICA

8.1 Middle East & Africa Superconducting MRI Equipment by Country

8.1.1 Middle East & Africa Superconducting MRI Equipment Sales by Country (2021-2026)

8.1.2 Middle East & Africa Superconducting MRI Equipment Revenue by Country (2021-2026)

8.2 Middle East & Africa Superconducting MRI Equipment Sales by Type (2021-2026)

8.3 Middle East & Africa Superconducting MRI Equipment Sales by Application (2021-2026)

8.4 Egypt

8.5 South Africa

8.6 Israel

8.7 Turkey

8.8 GCC Countries

9 MARKET DRIVERS, CHALLENGES AND TRENDS

9.1 Market Drivers & Growth Opportunities

9.2 Market Challenges & Risks

9.3 Industry Trends

10 MANUFACTURING COST STRUCTURE ANALYSIS

10.1 Raw Material and Suppliers

10.2 Manufacturing Cost Structure Analysis of Superconducting MRI Equipment

10.3 Manufacturing Process Analysis of Superconducting MRI Equipment

10.4 Industry Chain Structure of Superconducting MRI Equipment

11 MARKETING, DISTRIBUTORS AND CUSTOMER

11.1 Sales Channel

11.1.1 Direct Channels

11.1.2 Indirect Channels

11.2 Superconducting MRI Equipment Distributors

11.3 Superconducting MRI Equipment Customer

12 WORLD FORECAST REVIEW FOR SUPERCONDUCTING MRI EQUIPMENT BY GEOGRAPHIC REGION

12.1 Global Superconducting MRI Equipment Market Size Forecast by Region

12.1.1 Global Superconducting MRI Equipment Forecast by Region (2027-2032)

12.1.2 Global Superconducting MRI Equipment Annual Revenue Forecast by Region (2027-2032)

12.2 Americas Forecast by Country (2027-2032)

12.3 APAC Forecast by Region (2027-2032)

12.4 Europe Forecast by Country (2027-2032)

12.5 Middle East & Africa Forecast by Country (2027-2032)

12.6 Global Superconducting MRI Equipment Forecast by Type (2027-2032)

12.7 Global Superconducting MRI Equipment Forecast by Application (2027-2032)

13 KEY PLAYERS ANALYSIS

13.1 Siemens

13.1.1 Siemens Company Information

13.1.2 Siemens Superconducting MRI Equipment Product Portfolios and Specifications

13.1.3 Siemens Superconducting MRI Equipment Sales, Revenue, Price and Gross Margin (2021-2026)

13.1.4 Siemens Main Business Overview

13.1.5 Siemens Latest Developments

13.2 Philips

13.2.1 Philips Company Information

13.2.2 Philips Superconducting MRI Equipment Product Portfolios and Specifications

13.2.3 Philips Superconducting MRI Equipment Sales, Revenue, Price and Gross Margin (2021-2026)

13.2.4 Philips Main Business Overview

13.2.5 Philips Latest Developments

13.3 GE HealthCare

13.3.1 GE HealthCare Company Information

13.3.2 GE HealthCare Superconducting MRI Equipment Product Portfolios and Specifications

13.3.3 GE HealthCare Superconducting MRI Equipment Sales, Revenue, Price and Gross Margin (2021-2026)

13.3.4 GE HealthCare Main Business Overview

13.3.5 GE HealthCare Latest Developments

13.4 Canon Medical

13.4.1 Canon Medical Company Information

13.4.2 Canon Medical Superconducting MRI Equipment Product Portfolios and Specifications

13.4.3 Canon Medical Superconducting MRI Equipment Sales, Revenue, Price and Gross Margin (2021-2026)

13.4.4 Canon Medical Main Business Overview

13.4.5 Canon Medical Latest Developments

13.5 FUJIFILM

13.5.1 FUJIFILM Company Information

13.5.2 FUJIFILM Superconducting MRI Equipment Product Portfolios and Specifications

13.5.3 FUJIFILM Superconducting MRI Equipment Sales, Revenue, Price and Gross

Margin (2021-2026)

13.5.4 FUJIFILM Main Business Overview

13.5.5 FUJIFILM Latest Developments

13.6 United Imaging

13.6.1 United Imaging Company Information

13.6.2 United Imaging Superconducting MRI Equipment Product Portfolios and Specifications

13.6.3 United Imaging Superconducting MRI Equipment Sales, Revenue, Price and Gross Margin (2021-2026)

13.6.4 United Imaging Main Business Overview

13.6.5 United Imaging Latest Developments

13.7 Neusoft Medical

13.7.1 Neusoft Medical Company Information

13.7.2 Neusoft Medical Superconducting MRI Equipment Product Portfolios and Specifications

13.7.3 Neusoft Medical Superconducting MRI Equipment Sales, Revenue, Price and Gross Margin (2021-2026)

13.7.4 Neusoft Medical Main Business Overview

13.7.5 Neusoft Medical Latest Developments

13.8 XGY

13.8.1 XGY Company Information

13.8.2 XGY Superconducting MRI Equipment Product Portfolios and Specifications

13.8.3 XGY Superconducting MRI Equipment Sales, Revenue, Price and Gross Margin (2021-2026)

13.8.4 XGY Main Business Overview

13.8.5 XGY Latest Developments

13.9 Anke

13.9.1 Anke Company Information

13.9.2 Anke Superconducting MRI Equipment Product Portfolios and Specifications

13.9.3 Anke Superconducting MRI Equipment Sales, Revenue, Price and Gross Margin (2021-2026)

13.9.4 Anke Main Business Overview

13.9.5 Anke Latest Developments

13.10 Alltech

13.10.1 Alltech Company Information

13.10.2 Alltech Superconducting MRI Equipment Product Portfolios and Specifications

13.10.3 Alltech Superconducting MRI Equipment Sales, Revenue, Price and Gross Margin (2021-2026)

13.10.4 Alltech Main Business Overview

13.10.5 Alltech Latest Developments

13.11 Wandong

13.11.1 Wandong Company Information

13.11.2 Wandong Superconducting MRI Equipment Product Portfolios and Specifications

13.11.3 Wandong Superconducting MRI Equipment Sales, Revenue, Price and Gross Margin (2021-2026)

13.11.4 Wandong Main Business Overview

13.11.5 Wandong Latest Developments

13.12 BASDA

13.12.1 BASDA Company Information

13.12.2 BASDA Superconducting MRI Equipment Product Portfolios and Specifications

13.12.3 BASDA Superconducting MRI Equipment Sales, Revenue, Price and Gross Margin (2021-2026)

13.12.4 BASDA Main Business Overview

13.12.5 BASDA Latest Developments

13.13 Esaote S.p.A.

13.13.1 Esaote S.p.A. Company Information

13.13.2 Esaote S.p.A. Superconducting MRI Equipment Product Portfolios and Specifications

13.13.3 Esaote S.p.A. Superconducting MRI Equipment Sales, Revenue, Price and Gross Margin (2021-2026)

13.13.4 Esaote S.p.A. Main Business Overview

13.13.5 Esaote S.p.A. Latest Developments

14 RESEARCH FINDINGS AND CONCLUSION

List Of Tables

LIST OF TABLES

Table 1. Superconducting MRI Equipment Annual Sales CAGR by Geographic Region (2021, 2025 & 2032) & (\$ millions)

Table 2. Superconducting MRI Equipment Annual Sales CAGR by Country/Region (2021, 2025 & 2032) & (\$ millions)

Table 3. Major Players of 1.5T

Table 4. Major Players of 3.0T

Table 5. Major Players of Others

Table 6. Global Superconducting MRI Equipment Sales by Type (2021-2026) & (Units)

Table 7. Global Superconducting MRI Equipment Sales Market Share by Type (2021-2026)

Table 8. Global Superconducting MRI Equipment Revenue by Type (2021-2026) & (\$ million)

Table 9. Global Superconducting MRI Equipment Revenue Market Share by Type (2021-2026)

Table 10. Global Superconducting MRI Equipment Sale Price by Type (2021-2026) & (K US\$/Unit)

Table 11. Major Players of Closed-Bore Superconducting MRI Systems

Table 12. Major Players of Open-Bore Superconducting MRI Systems

Table 13. Global Superconducting MRI Equipment Sales by Pore ??Structure (2021-2026) & (Units)

Table 14. Global Superconducting MRI Equipment Sales Market Share by Pore ??Structure (2021-2026)

Table 15. Global Superconducting MRI Equipment Revenue by Pore ??Structure (2021-2026) & (\$ million)

Table 16. Global Superconducting MRI Equipment Revenue Market Share by Pore ??Structure (2021-2026)

Table 17. Global Superconducting MRI Equipment Sale Price by Pore ??Structure (2021-2026) & (K US\$/Unit)

Table 18. Major Players of Fixed Superconducting MRI Systems

Table 19. Major Players of Mobile Superconducting MRI Systems

Table 20. Global Superconducting MRI Equipment Sales by Installation and Mobility (2021-2026) & (Units)

Table 21. Global Superconducting MRI Equipment Sales Market Share by Installation and Mobility (2021-2026)

Table 22. Global Superconducting MRI Equipment Revenue by Installation and Mobility

(2021-2026) & (\$ million)

Table 23. Global Superconducting MRI Equipment Revenue Market Share by Installation and Mobility (2021-2026)

Table 24. Global Superconducting MRI Equipment Sale Price by Installation and Mobility (2021-2026) & (K US\$/Unit)

Table 25. Global Superconducting MRI Equipment Sale by Application (2021-2026) & (Units)

Table 26. Global Superconducting MRI Equipment Sale Market Share by Application (2021-2026)

Table 27. Global Superconducting MRI Equipment Revenue by Application (2021-2026) & (\$ million)

Table 28. Global Superconducting MRI Equipment Revenue Market Share by Application (2021-2026)

Table 29. Global Superconducting MRI Equipment Sale Price by Application (2021-2026) & (K US\$/Unit)

Table 30. Global Superconducting MRI Equipment Sales by Company (2021-2026) & (Units)

Table 31. Global Superconducting MRI Equipment Sales Market Share by Company (2021-2026)

Table 32. Global Superconducting MRI Equipment Revenue by Company (2021-2026) & (\$ millions)

Table 33. Global Superconducting MRI Equipment Revenue Market Share by Company (2021-2026)

Table 34. Global Superconducting MRI Equipment Sale Price by Company (2021-2026) & (K US\$/Unit)

Table 35. Key Manufacturers Superconducting MRI Equipment Producing Area Distribution and Sales Area

Table 36. Players Superconducting MRI Equipment Products Offered

Table 37. Superconducting MRI Equipment Concentration Ratio (CR3, CR5 and CR10) & (2024-2026)

Table 38. New Products and Potential Entrants

Table 39. Market M&A Activity & Strategy

Table 40. Global Superconducting MRI Equipment Sales by Geographic Region (2021-2026) & (Units)

Table 41. Global Superconducting MRI Equipment Sales Market Share Geographic Region (2021-2026)

Table 42. Global Superconducting MRI Equipment Revenue by Geographic Region (2021-2026) & (\$ millions)

Table 43. Global Superconducting MRI Equipment Revenue Market Share by

Geographic Region (2021-2026)

Table 44. Global Superconducting MRI Equipment Sales by Country/Region (2021-2026) & (Units)

Table 45. Global Superconducting MRI Equipment Sales Market Share by Country/Region (2021-2026)

Table 46. Global Superconducting MRI Equipment Revenue by Country/Region (2021-2026) & (\$ millions)

Table 47. Global Superconducting MRI Equipment Revenue Market Share by Country/Region (2021-2026)

Table 48. Americas Superconducting MRI Equipment Sales by Country (2021-2026) & (Units)

Table 49. Americas Superconducting MRI Equipment Sales Market Share by Country (2021-2026)

Table 50. Americas Superconducting MRI Equipment Revenue by Country (2021-2026) & (\$ millions)

Table 51. Americas Superconducting MRI Equipment Sales by Type (2021-2026) & (Units)

Table 52. Americas Superconducting MRI Equipment Sales by Application (2021-2026) & (Units)

Table 53. APAC Superconducting MRI Equipment Sales by Region (2021-2026) & (Units)

Table 54. APAC Superconducting MRI Equipment Sales Market Share by Region (2021-2026)

Table 55. APAC Superconducting MRI Equipment Revenue by Region (2021-2026) & (\$ millions)

Table 56. APAC Superconducting MRI Equipment Sales by Type (2021-2026) & (Units)

Table 57. APAC Superconducting MRI Equipment Sales by Application (2021-2026) & (Units)

Table 58. Europe Superconducting MRI Equipment Sales by Country (2021-2026) & (Units)

Table 59. Europe Superconducting MRI Equipment Revenue by Country (2021-2026) & (\$ millions)

Table 60. Europe Superconducting MRI Equipment Sales by Type (2021-2026) & (Units)

Table 61. Europe Superconducting MRI Equipment Sales by Application (2021-2026) & (Units)

Table 62. Middle East & Africa Superconducting MRI Equipment Sales by Country (2021-2026) & (Units)

Table 63. Middle East & Africa Superconducting MRI Equipment Revenue Market Share

by Country (2021-2026)

Table 64. Middle East & Africa Superconducting MRI Equipment Sales by Type (2021-2026) & (Units)

Table 65. Middle East & Africa Superconducting MRI Equipment Sales by Application (2021-2026) & (Units)

Table 66. Key Market Drivers & Growth Opportunities of Superconducting MRI Equipment

Table 67. Key Market Challenges & Risks of Superconducting MRI Equipment

Table 68. Key Industry Trends of Superconducting MRI Equipment

Table 69. Superconducting MRI Equipment Raw Material

Table 70. Key Suppliers of Raw Materials

Table 71. Superconducting MRI Equipment Distributors List

Table 72. Superconducting MRI Equipment Customer List

Table 73. Global Superconducting MRI Equipment Sales Forecast by Region (2027-2032) & (Units)

Table 74. Global Superconducting MRI Equipment Revenue Forecast by Region (2027-2032) & (\$ millions)

Table 75. Americas Superconducting MRI Equipment Sales Forecast by Country (2027-2032) & (Units)

Table 76. Americas Superconducting MRI Equipment Annual Revenue Forecast by Country (2027-2032) & (\$ millions)

Table 77. APAC Superconducting MRI Equipment Sales Forecast by Region (2027-2032) & (Units)

Table 78. APAC Superconducting MRI Equipment Annual Revenue Forecast by Region (2027-2032) & (\$ millions)

Table 79. Europe Superconducting MRI Equipment Sales Forecast by Country (2027-2032) & (Units)

Table 80. Europe Superconducting MRI Equipment Revenue Forecast by Country (2027-2032) & (\$ millions)

Table 81. Middle East & Africa Superconducting MRI Equipment Sales Forecast by Country (2027-2032) & (Units)

Table 82. Middle East & Africa Superconducting MRI Equipment Revenue Forecast by Country (2027-2032) & (\$ millions)

Table 83. Global Superconducting MRI Equipment Sales Forecast by Type (2027-2032) & (Units)

Table 84. Global Superconducting MRI Equipment Revenue Forecast by Type (2027-2032) & (\$ millions)

Table 85. Global Superconducting MRI Equipment Sales Forecast by Application (2027-2032) & (Units)

Table 86. Global Superconducting MRI Equipment Revenue Forecast by Application (2027-2032) & (\$ millions)

Table 87. Siemens Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 88. Siemens Superconducting MRI Equipment Product Portfolios and Specifications

Table 89. Siemens Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 90. Siemens Main Business

Table 91. Siemens Latest Developments

Table 92. Philips Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 93. Philips Superconducting MRI Equipment Product Portfolios and Specifications

Table 94. Philips Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 95. Philips Main Business

Table 96. Philips Latest Developments

Table 97. GE HealthCare Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 98. GE HealthCare Superconducting MRI Equipment Product Portfolios and Specifications

Table 99. GE HealthCare Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 100. GE HealthCare Main Business

Table 101. GE HealthCare Latest Developments

Table 102. Canon Medical Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 103. Canon Medical Superconducting MRI Equipment Product Portfolios and Specifications

Table 104. Canon Medical Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 105. Canon Medical Main Business

Table 106. Canon Medical Latest Developments

Table 107. FUJIFILM Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 108. FUJIFILM Superconducting MRI Equipment Product Portfolios and Specifications

Table 109. FUJIFILM Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 110. FUJIFILM Main Business

Table 111. FUJIFILM Latest Developments

Table 112. United Imaging Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 113. United Imaging Superconducting MRI Equipment Product Portfolios and Specifications

Table 114. United Imaging Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 115. United Imaging Main Business

Table 116. United Imaging Latest Developments

Table 117. Neusoft Medical Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 118. Neusoft Medical Superconducting MRI Equipment Product Portfolios and Specifications

Table 119. Neusoft Medical Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 120. Neusoft Medical Main Business

Table 121. Neusoft Medical Latest Developments

Table 122. XGY Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 123. XGY Superconducting MRI Equipment Product Portfolios and Specifications

Table 124. XGY Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 125. XGY Main Business

Table 126. XGY Latest Developments

Table 127. Anke Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 128. Anke Superconducting MRI Equipment Product Portfolios and Specifications

Table 129. Anke Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 130. Anke Main Business

Table 131. Anke Latest Developments

Table 132. Alltech Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 133. Alltech Superconducting MRI Equipment Product Portfolios and Specifications

Table 134. Alltech Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 135. Alltech Main Business

Table 136. Alltech Latest Developments

Table 137. Wandong Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 138. Wandong Superconducting MRI Equipment Product Portfolios and Specifications

Table 139. Wandong Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 140. Wandong Main Business

Table 141. Wandong Latest Developments

Table 142. BASDA Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 143. BASDA Superconducting MRI Equipment Product Portfolios and Specifications

Table 144. BASDA Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 145. BASDA Main Business

Table 146. BASDA Latest Developments

Table 147. Esaote S.p.A. Basic Information, Superconducting MRI Equipment Manufacturing Base, Sales Area and Its Competitors

Table 148. Esaote S.p.A. Superconducting MRI Equipment Product Portfolios and Specifications

Table 149. Esaote S.p.A. Superconducting MRI Equipment Sales (Units), Revenue (\$ Million), Price (K US\$/Unit) and Gross Margin (2021-2026)

Table 150. Esaote S.p.A. Main Business

Table 151. Esaote S.p.A. Latest Developments

List Of Figures

LIST OF FIGURES

- Figure 1. Picture of Superconducting MRI Equipment
- Figure 2. Superconducting MRI Equipment Report Years Considered
- Figure 3. Research Objectives
- Figure 4. Research Methodology
- Figure 5. Research Process and Data Source
- Figure 6. Global Superconducting MRI Equipment Sales Growth Rate 2021-2032 (Units)
- Figure 7. Global Superconducting MRI Equipment Revenue Growth Rate 2021-2032 (\$ millions)
- Figure 8. Superconducting MRI Equipment Sales by Geographic Region (2021, 2025 & 2032) & (\$ millions)
- Figure 9. Superconducting MRI Equipment Sales Market Share by Country/Region (2025)
- Figure 10. Superconducting MRI Equipment Sales Market Share by Country/Region (2021, 2025 & 2032)
- Figure 11. Product Picture of 1.5T
- Figure 12. Product Picture of 3.0T
- Figure 13. Product Picture of Others
- Figure 14. Global Superconducting MRI Equipment Sales Market Share by Type in 2026
- Figure 15. Global Superconducting MRI Equipment Revenue Market Share by Type (2021-2026)
- Figure 16. Product Picture of Closed-Bore Superconducting MRI Systems
- Figure 17. Product Picture of Open-Bore Superconducting MRI Systems
- Figure 18. Global Superconducting MRI Equipment Sales Market Share by Pore ??Structure in 2026
- Figure 19. Global Superconducting MRI Equipment Revenue Market Share by Pore ??Structure (2021-2026)
- Figure 20. Product Picture of Fixed Superconducting MRI Systems
- Figure 21. Product Picture of Mobile Superconducting MRI Systems
- Figure 22. Global Superconducting MRI Equipment Sales Market Share by Installation and Mobility in 2026
- Figure 23. Global Superconducting MRI Equipment Revenue Market Share by Installation and Mobility (2021-2026)
- Figure 24. Superconducting MRI Equipment Consumed in Hospitals

Figure 25. Global Superconducting MRI Equipment Market: Hospitals (2021-2026) & (Units)

Figure 26. Superconducting MRI Equipment Consumed in Imaging Centers

Figure 27. Global Superconducting MRI Equipment Market: Imaging Centers (2021-2026) & (Units)

Figure 28. Superconducting MRI Equipment Consumed in Other

Figure 29. Global Superconducting MRI Equipment Market: Other (2021-2026) & (Units)

Figure 30. Global Superconducting MRI Equipment Sale Market Share by Application (2025)

Figure 31. Global Superconducting MRI Equipment Revenue Market Share by Application in 2025

Figure 32. Superconducting MRI Equipment Sales by Company in 2025 (Units)

Figure 33. Global Superconducting MRI Equipment Sales Market Share by Company in 2025

Figure 34. Superconducting MRI Equipment Revenue by Company in 2025 (\$ millions)

Figure 35. Global Superconducting MRI Equipment Revenue Market Share by Company in 2025

Figure 36. Global Superconducting MRI Equipment Sales Market Share by Geographic Region (2021-2026)

Figure 37. Global Superconducting MRI Equipment Revenue Market Share by Geographic Region in 2025

Figure 38. Americas Superconducting MRI Equipment Sales 2021-2026 (Units)

Figure 39. Americas Superconducting MRI Equipment Revenue 2021-2026 (\$ millions)

Figure 40. APAC Superconducting MRI Equipment Sales 2021-2026 (Units)

Figure 41. APAC Superconducting MRI Equipment Revenue 2021-2026 (\$ millions)

Figure 42. Europe Superconducting MRI Equipment Sales 2021-2026 (Units)

Figure 43. Europe Superconducting MRI Equipment Revenue 2021-2026 (\$ millions)

Figure 44. Middle East & Africa Superconducting MRI Equipment Sales 2021-2026 (Units)

Figure 45. Middle East & Africa Superconducting MRI Equipment Revenue 2021-2026 (\$ millions)

Figure 46. Americas Superconducting MRI Equipment Sales Market Share by Country in 2025

Figure 47. Americas Superconducting MRI Equipment Revenue Market Share by Country (2021-2026)

Figure 48. Americas Superconducting MRI Equipment Sales Market Share by Type (2021-2026)

Figure 49. Americas Superconducting MRI Equipment Sales Market Share by Application (2021-2026)

Figure 50. United States Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 51. Canada Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 52. Mexico Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 53. Brazil Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 54. APAC Superconducting MRI Equipment Sales Market Share by Region in 2025

Figure 55. APAC Superconducting MRI Equipment Revenue Market Share by Region (2021-2026)

Figure 56. APAC Superconducting MRI Equipment Sales Market Share by Type (2021-2026)

Figure 57. APAC Superconducting MRI Equipment Sales Market Share by Application (2021-2026)

Figure 58. China Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 59. Japan Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 60. South Korea Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 61. Southeast Asia Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 62. India Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 63. Australia Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 64. China Taiwan Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 65. Europe Superconducting MRI Equipment Sales Market Share by Country in 2025

Figure 66. Europe Superconducting MRI Equipment Revenue Market Share by Country (2021-2026)

Figure 67. Europe Superconducting MRI Equipment Sales Market Share by Type (2021-2026)

Figure 68. Europe Superconducting MRI Equipment Sales Market Share by Application (2021-2026)

Figure 69. Germany Superconducting MRI Equipment Revenue Growth 2021-2026 (\$

millions)

Figure 70. France Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 71. UK Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 72. Italy Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 73. Russia Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 74. Middle East & Africa Superconducting MRI Equipment Sales Market Share by Country (2021-2026)

Figure 75. Middle East & Africa Superconducting MRI Equipment Sales Market Share by Type (2021-2026)

Figure 76. Middle East & Africa Superconducting MRI Equipment Sales Market Share by Application (2021-2026)

Figure 77. Egypt Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 78. South Africa Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 79. Israel Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 80. Turkey Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 81. GCC Countries Superconducting MRI Equipment Revenue Growth 2021-2026 (\$ millions)

Figure 82. Manufacturing Cost Structure Analysis of Superconducting MRI Equipment in 2026

Figure 83. Manufacturing Process Analysis of Superconducting MRI Equipment

Figure 84. Industry Chain Structure of Superconducting MRI Equipment

Figure 85. Channels of Distribution

Figure 86. Global Superconducting MRI Equipment Sales Market Forecast by Region (2027-2032)

Figure 87. Global Superconducting MRI Equipment Revenue Market Share Forecast by Region (2027-2032)

Figure 88. Global Superconducting MRI Equipment Sales Market Share Forecast by Type (2027-2032)

Figure 89. Global Superconducting MRI Equipment Revenue Market Share Forecast by Type (2027-2032)

Figure 90. Global Superconducting MRI Equipment Sales Market Share Forecast by

Application (2027-2032)

Figure 91. Global Superconducting MRI Equipment Revenue Market Share Forecast by Application (2027-2032)

I would like to order

Product name: Global Superconducting MRI Equipment Market Growth 2026-2032

Product link: <https://marketpublishers.com/r/S3BCC6C3FE4BEN.html>

Price: US\$ 3,660.00 (Single User License / Electronic Delivery)

If you want to order Corporate License or Hard Copy, please, contact our Customer Service:

info@marketpublishers.com

Payment

To pay by Credit Card (Visa, MasterCard, American Express, PayPal), please, click button on product page <https://marketpublishers.com/r/S3BCC6C3FE4BEN.html>