

Medical Device Reprocessing Market - Global Industry Size, Share, Trends, Opportunity, and Forecast, Segmented By Type (Reprocessing Support and Services, Reprocessed Medical Devices), By Devices Type (Critical Devices, Semi-Critical Devices, Non-Critical Devices), By Application (Cardiology, Gastroenterology, Gynecology, Arthroscopy & Orthopedic Surgery, General Surgery, Others), By Region, and By Competition, 2019-2029F

<https://marketpublishers.com/r/MB8A19F5558AEN.html>

Date: May 2024

Pages: 188

Price: US\$ 4,500.00 (Single User License)

ID: MB8A19F5558AEN

Abstracts

Global Medical Device Reprocessing Market was valued at USD 2.11 billion in 2023 and experience a steady growth in the forecast period at a CAGR of 13.93% through 2029. Medical device reprocessing refers to the process of cleaning, disinfecting, sterilizing, and refurbishing medical devices for safe and effective reuse in healthcare settings. Reprocessing allows healthcare facilities to extend the lifecycle of medical devices, reduce costs, minimize waste, and promote environmental sustainability. Used medical devices are collected and sorted based on their type, material composition, and reprocessing requirements. Devices may include surgical instruments, endoscopes, catheters, and other reusable medical equipment. The first step in reprocessing involves thorough cleaning of the medical devices to remove organic and inorganic contaminants, such as blood, tissue, and debris.

Cleaning may involve manual or automated techniques, including mechanical scrubbing, ultrasonic cleaning, and enzymatic detergents, to ensure the removal of visible and invisible residues. Following cleaning, medical devices undergo disinfection to eliminate microorganisms and reduce the risk of infection.

transmission. Disinfection methods may include chemical disinfection, such as immersion in high-level disinfectants or exposure to disinfectant solutions, to achieve a high level of microbial kill without compromising device integrity. Sterilization is a critical step in the reprocessing process to ensure the complete elimination of microorganisms, including bacteria, viruses, and spores, from the surface of medical devices. Sterilization methods may include steam sterilization (autoclaving), ethylene oxide (ETO) sterilization, hydrogen peroxide gas plasma sterilization, and radiation sterilization, depending on the device type and material compatibility.

Advances in reprocessing technologies have led to the development of more efficient and effective methods for cleaning, sterilizing, and refurbishing medical devices. Automation, robotics, and advanced sterilization techniques have improved the safety and quality of reprocessed devices, driving market growth. As the benefits of medical device reprocessing become more widely recognized within the healthcare industry, there is increasing awareness and acceptance of reprocessed devices among healthcare providers and institutions. This growing acceptance drives market growth as more healthcare facilities incorporate reprocessed devices into their procurement practices.

The medical device reprocessing market is experiencing consolidation as larger companies acquire smaller players to expand their service offerings and geographic presence. This consolidation drives market growth by increasing access to reprocessing services and promoting economies of scale. The rising cost of healthcare services, coupled with budget constraints faced by healthcare providers, has prompted healthcare facilities to explore cost-saving measures such as medical device reprocessing. Reprocessing allows healthcare facilities to optimize resource utilization and allocate funds to other areas of patient care.

Key Market Drivers

Technological Advancements

Automation and robotics have revolutionized medical device reprocessing by streamlining workflows, reducing human error, and improving efficiency. Automated systems can handle tasks such as cleaning, disinfection, and sterilization with precision and consistency, leading to higher quality outcomes. Technological innovations have led to the development of advanced sterilization techniques that ensure the thorough disinfection and sterilization of medical devices. These techniques may include low-temperature sterilization methods such as hydrogen peroxide gas plasma, ozone

sterilization, and vaporized hydrogen peroxide, which are effective for heat-sensitive devices.

Biological indicators and monitoring systems play a crucial role in verifying the effectiveness of sterilization processes. Advanced monitoring systems can track sterilization parameters such as temperature, pressure, and exposure time in real-time, ensuring compliance with regulatory standards and providing assurance of sterility. Technological advancements have enabled the safe and effective reprocessing of single-use medical devices (SUDs). Reprocessing technologies for SUDs may involve sophisticated cleaning methods, compatibility testing, and validation processes to ensure the devices can be safely reused without compromising patient safety.

Advanced data management and traceability systems enable healthcare facilities to track the entire lifecycle of reprocessed medical devices. These systems provide comprehensive documentation of reprocessing cycles, including cleaning and sterilization parameters, maintenance records, and device usage history, ensuring accountability and compliance with regulatory requirements. Technological advancements have improved compatibility testing and validation processes for reprocessed medical devices.

Advanced testing methods assess the compatibility of devices with reprocessing chemicals, sterilization methods, and packaging materials, ensuring that devices remain safe and functional throughout the reprocessing cycle. Integration with healthcare information systems allows for seamless communication and data exchange between reprocessing facilities and healthcare providers. Electronic tracking systems, barcoding technology, and RFID (Radio Frequency Identification) tags enable efficient inventory management, tracking of device usage, and timely reordering of reprocessed devices. This factor will help in the development of the Global Medical Device Reprocessing Market.

Growing Awareness and Acceptance

Healthcare facilities are increasingly pressured to reduce costs while maintaining high-quality patient care. Medical device reprocessing offers a cost-effective alternative to purchasing new devices, allowing healthcare providers to achieve significant savings without compromising patient safety or quality of care. There is a growing emphasis on sustainability within the healthcare industry. Reprocessing medical devices helps reduce medical waste and minimize the environmental impact of healthcare operations, aligning with sustainability initiatives and promoting a more environmentally

conscious approach to healthcare delivery. Regulatory bodies, such as the FDA (Food and Drug Administration) in the United States and the European Medicines Agency (EMA) in the European Union, have established guidelines and standards for the reprocessing of medical devices to ensure patient safety and quality of care.

Compliance with these regulations is driving the adoption of safe and effective reprocessing practices. Advances in reprocessing technologies have improved the safety, efficiency, and effectiveness of the reprocessing process. Automation, robotics, and advanced sterilization techniques have enhanced the quality assurance and reliability of reprocessed devices, contributing to their growing acceptance among healthcare providers. As the body of evidence supporting the safety and efficacy of reprocessed medical devices continues to grow, healthcare providers are becoming more confident in incorporating reprocessed devices into their clinical practice. Studies and research demonstrating the equivalent performance of reprocessed devices compared to new devices have contributed to increased acceptance and adoption.

Collaboration among industry stakeholders, including healthcare providers, manufacturers, regulatory agencies, and reprocessing companies, has played a crucial role in raising awareness and promoting the acceptance of medical device reprocessing. Educational initiatives, training programs, and industry conferences help disseminate information about reprocessing best practices and safety guidelines. Patients are becoming more aware of healthcare costs and environmental sustainability issues. As patients become more informed about reprocessing practices, they may express preferences for healthcare facilities that prioritize cost-effective and environmentally sustainable approaches to medical device management, driving demand for reprocessing services. This factor will pace up the demand of the Global Medical Device Reprocessing Market.

Rising preference of Environmental Sustainability

Reprocessing medical devices allows healthcare facilities to reuse devices multiple times instead of disposing of them after a single use. This reduces the volume of medical waste generated by hospitals and clinics, contributing to waste reduction efforts and minimizing the environmental impact of healthcare operations. Reprocessing medical devices conserves valuable resources, including raw materials, energy, and water, that would otherwise be used in the manufacturing of new devices. By extending the lifecycle of medical devices through reprocessing, healthcare facilities reduce the overall consumption of resources and promote resource conservation. The production

and disposal of medical devices can result in pollution and emissions that harm the environment and human health. Reprocessing medical devices reduces the need for manufacturing new devices, thereby mitigating pollution and emissions associated with the production process and reducing the environmental burden.

Medical device reprocessing contributes to the principles of a circular economy by promoting the reuse and recycling of resources. Instead of following a linear model of production and disposal, reprocessing establishes a closed-loop system where medical devices are reused, refurbished, and reintroduced into the healthcare system, reducing the need for new materials and minimizing waste generation. Many healthcare facilities are adopting corporate social responsibility initiatives that prioritize environmental sustainability and social impact. Incorporating medical device reprocessing into healthcare operations demonstrates a commitment to sustainable practices and environmental stewardship, enhancing the organization's reputation and fostering goodwill within the community.

Regulatory agencies increasingly emphasize environmental sustainability and waste reduction in healthcare settings. Compliance with environmental regulations and standards encourages healthcare facilities to adopt reprocessing practices as part of their environmental management strategies, driving the demand for reprocessing services and solutions. Patients and the general public are becoming more aware of environmental issues and sustainability concerns. As patients seek healthcare providers and facilities that prioritize environmental sustainability, there is growing pressure on healthcare organizations to adopt environmentally friendly practices, including medical device reprocessing. This factor will accelerate the demand of the Global Medical Device Reprocessing Market.

Key Market Challenges

Risk of Infection Transmission

Medical devices are not thoroughly cleaned and disinfected during the reprocessing process, residual organic material and pathogens may remain on the surfaces of the devices. This increases the risk of infection transmission to subsequent patients who come into contact with the inadequately reprocessed devices. Many medical devices, particularly those used in surgical procedures, are complex in design and may have intricate features that are difficult to clean and disinfect effectively. The presence of small crevices, channels, and lumens can harbor microbial contamination and compromise the efficacy of the reprocessing process. Reprocessing practices may

vary among healthcare facilities and reprocessing facilities, leading to inconsistencies in the quality and effectiveness of the reprocessing process. Variability in reprocessing practices increases the risk of inadequate cleaning and disinfection, potentially resulting in infection transmission to patients.

Some microorganisms, such as antibiotic-resistant bacteria (e.g., MRSA, VRE), are more resilient to disinfection and sterilization processes. If medical devices become contaminated with resistant microorganisms during patient use, there is a heightened risk of transmission to subsequent patients if the devices are not adequately reprocessed. Human error and procedural lapses during the reprocessing process can compromise the effectiveness of cleaning and disinfection procedures. Factors such as inadequate training, fatigue, distractions, and time constraints may contribute to errors and lapses in reprocessing practices, increasing the risk of infection transmission. Inadequate implementation of quality control measures, such as routine monitoring, testing, and validation of reprocessing processes, can contribute to the risk of infection transmission. Without robust quality control measures in place, healthcare facilities may fail to detect deficiencies in the reprocessing process that could compromise patient safety.

Public Perception and Acceptance

Despite the rigorous cleaning and sterilization processes involved in medical device reprocessing, some patients may perceive reprocessed devices as less safe compared to new devices. Concerns about the risk of infection transmission and cross-contamination may lead to skepticism and reluctance to accept reprocessed devices. Many patients and healthcare consumers may have limited awareness and understanding of medical device reprocessing and its safety protocols. Without sufficient education and information about the reprocessing process and regulatory standards, patients may harbor misconceptions and concerns about the safety and efficacy of reprocessed devices. Patients may inherently trust and prefer new medical devices over reprocessed ones due to perceptions of higher quality, reliability, and safety. Building trust and confidence in reprocessed devices requires transparent communication, evidence-based information, and assurances of adherence to rigorous reprocessing standards.

Negative media coverage of incidents or outbreaks related to medical device reprocessing can significantly impact public perception and acceptance of reprocessed devices. Sensationalized reporting of isolated incidents or regulatory violations may erode public trust and confidence in reprocessing practices. Cultural beliefs, societal

attitudes, and personal preferences may influence public perception and acceptance of reprocessed medical devices. Cultural biases or stigmas associated with reused products or equipment may influence patient preferences and choices regarding medical device use. Effective communication and transparency about the reprocessing process, safety protocols, and regulatory compliance are essential for addressing public concerns and building trust. Healthcare providers and reprocessing facilities must proactively communicate with patients, addressing their questions and concerns about reprocessed devices to promote understanding and acceptance.

Key Market Trends

Increasing Outsourcing of Reprocessing Services

Outsourcing reprocessing services to specialized third-party providers can offer cost savings compared to establishing and maintaining in-house reprocessing facilities. Outsourced reprocessing services may benefit from economies of scale, expertise, and operational efficiencies that result in lower costs for healthcare facilities. Healthcare providers increasingly prioritize their core competencies, such as patient care and clinical services, while outsourcing non-core activities, such as medical device reprocessing. Outsourcing reprocessing services allows healthcare facilities to allocate resources and personnel to essential clinical functions, improving operational focus and efficiency. Outsourced reprocessing providers often possess specialized expertise, experience, and resources dedicated to medical device reprocessing. These providers may employ trained personnel, utilize advanced reprocessing technologies, and implement stringent quality control measures to ensure the safety and efficacy of reprocessed devices.

Outsourced reprocessing services offer flexibility and scalability to accommodate fluctuations in demand, procedural volumes, and resource requirements. Healthcare facilities can adjust reprocessing services based on changing needs without the constraints of maintaining fixed infrastructure and staffing levels. Outsourcing reprocessing services may help healthcare facilities mitigate risks associated with medical device reprocessing, including infection transmission, regulatory non-compliance, and liability concerns. Third-party reprocessing providers assume responsibility for compliance, quality assurance, and risk management, reducing the burden on healthcare facilities. The medical device reprocessing market is experiencing consolidation, with specialized reprocessing providers emerging to meet the growing demand for outsourced reprocessing services. These specialized providers offer niche expertise, customized solutions, and value-added services tailored to the needs of

healthcare facilities.

Segmental Insights

Type Insights

Based on the type, the Reprocessed Medical Devices segment is projected to experience rapid growth in the Global Medical Device Reprocessing Market during the forecast period. Reprocessed medical devices offer a cost-effective alternative to purchasing new devices. Healthcare providers are under pressure to contain costs while maintaining high standards of patient care. Reprocessing allows them to achieve significant cost savings without compromising the quality or safety of medical devices. There is a growing emphasis on sustainability and environmental responsibility in healthcare. Reprocessing medical devices reduces medical waste and minimizes the environmental impact of healthcare operations.

Healthcare facilities are increasingly adopting reprocessed devices as part of their sustainability initiatives. The range of medical devices suitable for reprocessing has expanded significantly in recent years. Initially focused on simple, single-use devices, reprocessing now encompasses a wide variety of complex medical devices, including surgical instruments, endoscopes, and imaging equipment. This expansion of the reprocessed device portfolio has contributed to the growth of the segment. Healthcare providers are becoming more comfortable with the concept of reprocessed medical devices as they become more familiar with the process and its benefits. As the evidence supporting the safety and efficacy of reprocessed devices continues to accumulate, healthcare providers are more willing to incorporate reprocessed devices into their clinical practice.

Application Insights

Based on the application, the Arthroscopy & Orthopedic Surgery segment is projected to experience rapid growth in the Global Medical Device Reprocessing Market during the forecast period. There is a growing prevalence of orthopedic conditions and injuries worldwide, driven by factors such as aging populations, sports-related injuries, and lifestyle changes. As the demand for orthopedic surgeries, including arthroscopic procedures, rises, there is a corresponding increase in the need for reprocessed medical devices used in these surgeries. Healthcare expenditure on orthopedic surgeries and procedures continues to increase globally. However, healthcare providers face pressure to contain costs while maintaining high standards of patient

care. Reprocessing medical devices used in orthopedic surgery offers a cost-effective solution for healthcare facilities to manage expenses without compromising patient outcomes.

Advances in arthroscopic techniques and minimally invasive orthopedic procedures have expanded the scope of treatment options for various orthopedic conditions. As arthroscopy becomes more widely used in orthopedic surgery, there is a greater demand for reprocessed arthroscopic instruments and devices. Regulatory bodies impose strict regulations and standards for the reprocessing of medical devices, including those used in arthroscopy and orthopedic surgery. Compliance with these regulations ensures patient safety and reduces the risk of healthcare-associated infections (HAIs), driving the demand for reliable reprocessing solutions and services. Healthcare facilities are increasingly adopting sustainable practices, including the reprocessing of medical devices, to reduce waste and minimize environmental impact. Reprocessing arthroscopic and orthopedic surgical instruments aligns with these sustainability initiatives, contributing to the growth of the reprocessing market in this segment.

Regional Insights

Based on the region, the North America emerged as the dominant region in the Global Medical Device Reprocessing Market in 2023. In North America, specially the United States, has stringent regulatory standards for governing medical device reprocessing. Regulatory bodies such as the Food and Drug Administration (FDA) enforce rigorous requirements for the reprocessing of medical devices, ensuring safety and efficacy. The region boasts advanced healthcare infrastructure with a high prevalence of healthcare facilities and hospitals equipped with sophisticated medical devices. The need to optimize resources and control costs has led healthcare providers in North America to embrace medical device reprocessing as a cost-effective solution.

Rising healthcare costs and budget constraints have compelled healthcare providers to seek cost-effective alternatives to purchasing new medical devices. Reprocessing allows them to achieve significant cost savings without compromising patient care or safety. There is growing awareness and acceptance of medical device reprocessing among healthcare professionals and institutions in North America. As the benefits of reprocessing, such as cost savings and environmental sustainability, become more widely recognized, the adoption of reprocessing practices has increased.

Key Market Players

Stryker Corporation

Johnson & Johnson

Baxter International Inc

3M Company

Vanguard AG

Getinge AB

Olympus Corporation

Medtronic Plc.

Lumitos AG

Report Scope:

In this report, the Global Medical Device Reprocessing Market has been segmented into the following categories, in addition to the industry trends which have also been detailed below:

Medical Device Reprocessing Market, By Type:

Reprocessing Support and Services

Reprocessed Medical Devices

Medical Device Reprocessing Market, By Devices Type:

Critical Devices

Semi-Critical Devices

Non-Critical Devices

Medical Device Reprocessing Market, By Application:

Cardiology

Gastroenterology

Gynecology

Arthroscopy & Orthopedic Surgery

General Surgery

Others

Medical Device Reprocessing Market, By Region:

North America

United States

Canada

Mexico

Europe

Germany

United Kingdom

France

Italy

Spain

Asia-Pacific

China

Japan

India

Australia

South Korea

South America

Brazil

Argentina

Colombia

Middle East & Africa

South Africa

Saudi Arabia

UAE

Competitive Landscape

Company Profiles: Detailed analysis of the major companies present in the Global Medical Device Reprocessing Market.

Available Customizations:

Global Medical Device Reprocessing market report with the given market data, Tech Sci Research offers customizations according to a company's specific needs. The following customization options are available for the report:

Company Information

Detailed analysis and profiling of additional market players (up to five).

Contents

1. PRODUCT OVERVIEW

- 1.1. Market Definition
- 1.2. Scope of the Market
 - 1.2.1. Markets Covered
 - 1.2.2. Years Considered for Study
 - 1.2.3. Key Market Segmentations

2. RESEARCH METHODOLOGY

- 2.1. Objective of the Study
- 2.2. Baseline Methodology
- 2.3. Key Industry Partners
- 2.4. Major Association and Secondary Sources
- 2.5. Forecasting Methodology
- 2.6. Data Triangulation & Validation
- 2.7. Assumptions and Limitations

3. EXECUTIVE SUMMARY

- 3.1. Overview of the Market
- 3.2. Overview of Key Market Segmentations
- 3.3. Overview of Key Market Players
- 3.4. Overview of Key Regions/Countries
- 3.5. Overview of Market Drivers, Challenges, Trends

4. VOICE OF CUSTOMER

5. GLOBAL MEDICAL DEVICE REPROCESSING MARKET OUTLOOK

- 5.1. Market Size & Forecast
 - 5.1.1. By Value
- 5.2. Market Share & Forecast
 - 5.2.1. By Type (Reprocessing Support and Services, Reprocessed Medical Devices)
 - 5.2.2. By Devices Type (Critical Devices, Semi-Critical Devices, Non-Critical Devices)
 - 5.2.3. By Application (Cardiology, Gastroenterology, Gynecology, Arthroscopy &

Orthopedic Surgery, General Surgery, Others)

5.2.4. By Region

5.2.5. By Company (2023)

5.3. Market Map

6. NORTH AMERICA MEDICAL DEVICE REPROCESSING MARKET OUTLOOK

6.1. Market Size & Forecast

6.1.1. By Value

6.2. Market Share & Forecast

6.2.1. By Type

6.2.2. By Devices Type

6.2.3. By Application

6.2.4. By Country

6.3. North America: Country Analysis

6.3.1. United States Medical Device Reprocessing Market Outlook

6.3.1.1. Market Size & Forecast

6.3.1.1.1. By Value

6.3.1.2. Market Share & Forecast

6.3.1.2.1. By Type

6.3.1.2.2. By Devices Type

6.3.1.2.3. By Application

6.3.2. Canada Medical Device Reprocessing Market Outlook

6.3.2.1. Market Size & Forecast

6.3.2.1.1. By Value

6.3.2.2. Market Share & Forecast

6.3.2.2.1. By Type

6.3.2.2.2. By Devices Type

6.3.2.2.3. By Application

6.3.3. Mexico Medical Device Reprocessing Market Outlook

6.3.3.1. Market Size & Forecast

6.3.3.1.1. By Value

6.3.3.2. Market Share & Forecast

6.3.3.2.1. By Type

6.3.3.2.2. By Devices Type

6.3.3.2.3. By Application

7. EUROPE MEDICAL DEVICE REPROCESSING MARKET OUTLOOK

- 7.1. Market Size & Forecast
 - 7.1.1. By Value
- 7.2. Market Share & Forecast
 - 7.2.1. By Type
 - 7.2.2. By Devices Type
 - 7.2.3. By Application
 - 7.2.4. By Country
- 7.3. Europe: Country Analysis
 - 7.3.1. Germany Medical Device Reprocessing Market Outlook
 - 7.3.1.1. Market Size & Forecast
 - 7.3.1.1.1. By Value
 - 7.3.1.2. Market Share & Forecast
 - 7.3.1.2.1. By Type
 - 7.3.1.2.2. By Devices Type
 - 7.3.1.2.3. By Application
 - 7.3.2. United Kingdom Medical Device Reprocessing Market Outlook
 - 7.3.2.1. Market Size & Forecast
 - 7.3.2.1.1. By Value
 - 7.3.2.2. Market Share & Forecast
 - 7.3.2.2.1. By Type
 - 7.3.2.2.2. By Devices Type
 - 7.3.2.2.3. By Application
 - 7.3.3. Italy Medical Device Reprocessing Market Outlook
 - 7.3.3.1. Market Size & Forecast
 - 7.3.3.1.1. By Value
 - 7.3.3.2. Market Share & Forecast
 - 7.3.3.2.1. By Type
 - 7.3.3.2.2. By Devices Type
 - 7.3.3.2.3. By Application
 - 7.3.4. France Medical Device Reprocessing Market Outlook
 - 7.3.4.1. Market Size & Forecast
 - 7.3.4.1.1. By Value
 - 7.3.4.2. Market Share & Forecast
 - 7.3.4.2.1. By Type
 - 7.3.4.2.2. By Devices Type
 - 7.3.4.2.3. By Application
 - 7.3.5. Spain Medical Device Reprocessing Market Outlook
 - 7.3.5.1. Market Size & Forecast
 - 7.3.5.1.1. By Value

7.3.5.2. Market Share & Forecast

7.3.5.2.1. By Type

7.3.5.2.2. By Devices Type

7.3.5.2.3. By Application

8. ASIA-PACIFIC MEDICAL DEVICE REPROCESSING MARKET OUTLOOK

8.1. Market Size & Forecast

8.1.1. By Value

8.2. Market Share & Forecast

8.2.1. By Type

8.2.2. By Devices Type

8.2.3. By Application

8.2.4. By Country

8.3. Asia-Pacific: Country Analysis

8.3.1. China Medical Device Reprocessing Market Outlook

8.3.1.1. Market Size & Forecast

8.3.1.1.1. By Value

8.3.1.2. Market Share & Forecast

8.3.1.2.1. By Type

8.3.1.2.2. By Devices Type

8.3.1.2.3. By Application

8.3.2. India Medical Device Reprocessing Market Outlook

8.3.2.1. Market Size & Forecast

8.3.2.1.1. By Value

8.3.2.2. Market Share & Forecast

8.3.2.2.1. By Type

8.3.2.2.2. By Devices Type

8.3.2.2.3. By Application

8.3.3. Japan Medical Device Reprocessing Market Outlook

8.3.3.1. Market Size & Forecast

8.3.3.1.1. By Value

8.3.3.2. Market Share & Forecast

8.3.3.2.1. By Type

8.3.3.2.2. By Devices Type

8.3.3.2.3. By Application

8.3.4. South Korea Medical Device Reprocessing Market Outlook

8.3.4.1. Market Size & Forecast

8.3.4.1.1. By Value

- 8.3.4.2. Market Share & Forecast
 - 8.3.4.2.1. By Type
 - 8.3.4.2.2. By Devices Type
 - 8.3.4.2.3. By Application
- 8.3.5. Australia Medical Device Reprocessing Market Outlook
 - 8.3.5.1. Market Size & Forecast
 - 8.3.5.1.1. By Value
 - 8.3.5.2. Market Share & Forecast
 - 8.3.5.2.1. By Type
 - 8.3.5.2.2. By Devices Type
 - 8.3.5.2.3. By Application

9. SOUTH AMERICA MEDICAL DEVICE REPROCESSING MARKET OUTLOOK

- 9.1. Market Size & Forecast
 - 9.1.1. By Value
- 9.2. Market Share & Forecast
 - 9.2.1. By Type
 - 9.2.2. By Devices Type
 - 9.2.3. By Application
 - 9.2.4. By Country
- 9.3. South America: Country Analysis
 - 9.3.1. Brazil Medical Device Reprocessing Market Outlook
 - 9.3.1.1. Market Size & Forecast
 - 9.3.1.1.1. By Value
 - 9.3.1.2. Market Share & Forecast
 - 9.3.1.2.1. By Type
 - 9.3.1.2.2. By Devices Type
 - 9.3.1.2.3. By Application
 - 9.3.2. Argentina Medical Device Reprocessing Market Outlook
 - 9.3.2.1. Market Size & Forecast
 - 9.3.2.1.1. By Value
 - 9.3.2.2. Market Share & Forecast
 - 9.3.2.2.1. By Type
 - 9.3.2.2.2. By Devices Type
 - 9.3.2.2.3. By Application
 - 9.3.3. Colombia Medical Device Reprocessing Market Outlook
 - 9.3.3.1. Market Size & Forecast
 - 9.3.3.1.1. By Value

9.3.3.2. Market Share & Forecast

9.3.3.2.1. By Type

9.3.3.2.2. By Devices Type

9.3.3.2.3. By Application

10. MIDDLE EAST AND AFRICA MEDICAL DEVICE REPROCESSING MARKET OUTLOOK

10.1. Market Size & Forecast

10.1.1. By Value

10.2. Market Share & Forecast

10.2.1. By Type

10.2.2. By Devices Type

10.2.3. By Application

10.2.4. By Country

10.3. MEA: Country Analysis

10.3.1. South Africa Medical Device Reprocessing Market Outlook

10.3.1.1. Market Size & Forecast

10.3.1.1.1. By Value

10.3.1.2. Market Share & Forecast

10.3.1.2.1. By Type

10.3.1.2.2. By Devices Type

10.3.1.2.3. By Application

10.3.2. Saudi Arabia Medical Device Reprocessing Market Outlook

10.3.2.1. Market Size & Forecast

10.3.2.1.1. By Value

10.3.2.2. Market Share & Forecast

10.3.2.2.1. By Type

10.3.2.2.2. By Devices Type

10.3.2.2.3. By Application

10.3.3. UAE Medical Device Reprocessing Market Outlook

10.3.3.1. Market Size & Forecast

10.3.3.1.1. By Value

10.3.3.2. Market Share & Forecast

10.3.3.2.1. By Type

10.3.3.2.2. By Devices Type

10.3.3.2.3. By Application

11. MARKET DYNAMICS

11.1. Drivers

11.2. Challenges

12. MARKET TRENDS & DEVELOPMENTS

12.1. Merger & Acquisition (If Any)

12.2. Product Launches (If Any)

12.3. Recent Developments

13. PORTER'S FIVE FORCES ANALYSIS

13.1. Competition in the Industry

13.2. Potential of New Entrants

13.3. Power of Suppliers

13.4. Power of Customers

13.5. Threat of Substitute Product

14. COMPETITIVE LANDSCAPE

14.1. Stryker Corporation

14.1.1. Business Overview

14.1.2. Company Snapshot

14.1.3. Products & Services

14.1.4. Financials (As Reported)

14.1.5. Recent Developments

14.1.6. Key Personnel Details

14.1.7. SWOT Analysis

14.2. Johnson & Johnson

14.3. Baxter International Inc

14.4. 3M Company

14.5. Vanguard AG

14.6. Getinge AB

14.7. Olympus Corporation

14.8. Medtronic Plc.

14.9. Lumitos AG

15. STRATEGIC RECOMMENDATIONS

16. ABOUT US & DISCLAIMER

I would like to order

Product name: Medical Device Reprocessing Market - Global Industry Size, Share, Trends, Opportunity, and Forecast, Segmented By Type (Reprocessing Support and Services, Reprocessed Medical Devices), By Devices Type (Critical Devices, Semi-Critical Devices, Non-Critical Devices), By Application (Cardiology, Gastroenterology, Gynecology, Arthroscopy & Orthopedic Surgery, General Surgery, Others), By Region, and By Competition, 2019-2029F

Product link: <https://marketpublishers.com/r/MB8A19F5558AEN.html>

Price: US\$ 4,500.00 (Single User License / Electronic Delivery)

If you want to order Corporate License or Hard Copy, please, contact our Customer Service:

info@marketpublishers.com

Payment

To pay by Credit Card (Visa, MasterCard, American Express, PayPal), please, click button on product page <https://marketpublishers.com/r/MB8A19F5558AEN.html>