

Ultrasound Probe Disinfection Market - Global Industry Size, Share, Trends, Opportunity, and Forecast, Segmented By Products & Services (Instruments, Consumables, and Services), By Level of Disinfection (Low Level Disinfection, Intermediate Level Disinfection, and High Level Disinfection), By Type of Probe (Critical Devices, Noncritical Devices, and Semicritical Devices), By End User (Hospitals & Diagnostic Imaging Centers, Ambulatory Care Centers, Academic & Research Institutes, and Others), By Region, and By Competition, 2019-2029F

<https://marketpublishers.com/r/U68E95693C18EN.html>

Date: May 2024

Pages: 182

Price: US\$ 4,900.00 (Single User License)

ID: U68E95693C18EN

Abstracts

Global Ultrasound Probe Disinfection Market was valued at USD 523.64 million in 2023 and expected to reach a steady growth in the forecast period at a CAGR of 14.11% through 2029. Ultrasound probe disinfection is a critical process within healthcare settings, serving to mitigate the risk of healthcare-associated infections (HAIs) and uphold patient safety during ultrasound examinations and procedures. Ultrasound probes, pivotal medical devices used across various specialties such as obstetrics, gynecology, cardiology, urology, and radiology, require meticulous disinfection to eliminate or minimize the presence of pathogenic microorganisms, including bacteria, viruses, fungi, and spores.

Effective disinfection protocols are imperative to prevent the transmission of infectious agents between patients and healthcare providers, ensuring a hygienic healthcare environment. Healthcare facilities deploy diverse disinfection methods and

techniques to achieve optimal probe disinfection, encompassing chemical disinfection, high-level disinfection (HLD), ultraviolet (UV) light disinfection, hydrogen peroxide vapor disinfection, ozone-based disinfection, and automated disinfection systems. Proficiency in ultrasound probe disinfection procedures is essential for healthcare personnel, necessitating thorough training and education on proper disinfection techniques, equipment operation, safety precautions, and infection control protocols. Comprehensive training programs and ongoing education initiatives ensure adherence to standardized disinfection practices, fostering patient safety.

The escalating utilization of ultrasound imaging across medical specialties propels the demand for ultrasound probe disinfection solutions. As ultrasound procedures witness a global uptrend, healthcare facilities prioritize effective disinfection protocols to mitigate HAI risks and uphold patient safety. Advancements in disinfection technologies, including UV light, hydrogen peroxide vapor, ozone-based disinfection, and other innovative methods, drive market growth by offering more efficient, rapid, and environmentally sustainable disinfection solutions. Manufacturers respond to market demands by investing in research and development, aiming to create next-generation disinfection systems capable of achieving HLD while minimizing probe damage and enhancing workflow efficiency. These advancements underscore the evolving landscape of ultrasound probe disinfection, where innovation is pivotal in meeting the increasing demands for stringent infection control measures.

Ultrasound probe disinfection is integral to maintaining patient safety and reducing the incidence of HAIs in healthcare settings. Effective disinfection protocols, supported by ongoing advancements in technology and comprehensive training initiatives, ensure the integrity of ultrasound examinations and procedures across diverse medical specialties. As the demand for ultrasound imaging continues to rise, the market for ultrasound probe disinfection solutions is poised for sustained growth, driven by the imperative to uphold the highest standards of infection control and patient care.

Key Market Drivers

Increasing Adoption of Ultrasound Imaging

Ultrasound imaging is widely used across various medical specialties and clinical settings for diagnostic, therapeutic, and monitoring purposes. Its non-invasive nature, portability, and versatility make it a valuable tool in obstetrics, gynecology, cardiology, urology, radiology, and other medical fields. As the utilization of ultrasound imaging continues to expand globally, the number of ultrasound procedures performed

increases, driving the demand for ultrasound probes and, consequently, the need for effective disinfection solutions. Ultrasound probes come into direct contact with patients' skin or mucous membranes during imaging procedures. As a result, they can serve as potential vectors for the transmission of pathogens, including bacteria, viruses, and fungi. Healthcare facilities recognize the importance of implementing robust disinfection protocols to minimize the risk of healthcare-associated infections (HAIs) and ensure patient safety. The increasing adoption of ultrasound imaging underscores the need for effective disinfection practices to maintain a safe healthcare environment.

Regulatory agencies and professional organizations worldwide emphasize the importance of infection control measures, including the disinfection of medical devices such as ultrasound probes. Stringent regulatory requirements and infection control guidelines mandate the implementation of standardized disinfection protocols and the use of high-level disinfection (HLD) methods for semi-critical and critical devices. The increasing adoption of ultrasound imaging amplifies the importance of compliance with regulatory standards, driving the demand for advanced ultrasound probe disinfection solutions. The ultrasound probe disinfection market benefits from ongoing technological advancements aimed at enhancing disinfection efficacy, efficiency, and user-friendliness. Manufacturers develop innovative disinfection systems and technologies, such as automated disinfection machines, ultraviolet (UV) light systems, hydrogen peroxide vapor systems, and ozone-based disinfection solutions, to meet the evolving needs of healthcare facilities. The increasing adoption of ultrasound imaging encourages investment in advanced disinfection technologies to streamline disinfection workflows and ensure optimal patient care outcomes. This factor will help in the development of the Global Ultrasound Probe Disinfection Market.

Technological Advancements in Disinfection Technologies

Technological advancements have led to the development of disinfection technologies that offer enhanced efficacy in eliminating a wide range of pathogens from ultrasound probes. These advancements include ultraviolet (UV) light systems, hydrogen peroxide vapor systems, ozone-based disinfection solutions, and other innovative methods that effectively target and destroy harmful microorganisms, including bacteria, viruses, and fungi. Advanced disinfection technologies enable rapid disinfection processes, minimizing downtime between ultrasound procedures and optimizing workflow efficiency within healthcare settings. Automated disinfection machines equipped with advanced sensors, robotics, and intuitive user interfaces streamline the disinfection process, reducing the burden on healthcare staff and

ensuring consistent and reliable disinfection outcomes. Technological advancements in disinfection technologies address concerns regarding compatibility with different ultrasound probe materials and transducer designs. Manufacturers develop disinfection systems and disinfectants that are safe for use with a variety of probe materials, including plastics, metals, and elastomers, minimizing the risk of damage to probe components and ensuring long-term reliability and performance.

The shift towards environmentally sustainable disinfection solutions drives technological innovation in the ultrasound probe disinfection market. Manufacturers develop eco-friendly disinfection technologies that minimize the use of harmful chemicals, reduce water consumption, and lower overall environmental impact. Ozone-based disinfection systems, UV light systems, and other environmentally friendly disinfection methods align with healthcare facilities' sustainability goals while ensuring effective pathogen eradication. Advanced disinfection technologies feature user-friendly design elements and seamless integration into existing healthcare infrastructure. Intuitive interfaces, automated cycle programming, and real-time monitoring capabilities simplify the disinfection process and minimize the risk of human error.

Compact and ergonomic designs ensure compatibility with various examination rooms, operating suites, and point-of-care settings, enhancing versatility and usability. Technological advancements enable disinfection technologies to meet and exceed regulatory standards and guidelines for the disinfection of medical devices, including ultrasound probes. Manufacturers invest in research and development to ensure compliance with stringent regulatory requirements and obtain regulatory approvals for their disinfection products, providing healthcare facilities with confidence in the safety, efficacy, and quality of their disinfection solutions. This factor will pace up the demand of the Global Ultrasound Probe Disinfection Market.

Growing Healthcare Expenditure and Infrastructure Development

With growing healthcare expenditure, more individuals have access to healthcare services, leading to increased utilization of medical facilities, including those that offer ultrasound imaging services. As the number of ultrasound procedures and patient visits to healthcare facilities rises, the demand for effective disinfection of ultrasound probes also increases to ensure patient safety and prevent the transmission of infections. Infrastructure development initiatives, such as the construction of new hospitals, clinics, diagnostic centers, and outpatient facilities, contribute to the expansion of healthcare services in both urban and rural areas.

As healthcare facilities expand their capacity to accommodate a larger patient population, the need for reliable disinfection solutions for medical devices, including ultrasound probes, becomes more pronounced to maintain high standards of infection control and prevent healthcare-associated infections (HAIs). Healthcare providers prioritize infection prevention and control measures to mitigate the risk of HAIs and ensure patient safety. Ultrasound probes, which come into direct contact with patients during diagnostic procedures, represent potential vectors for the transmission of pathogens. As healthcare facilities invest in infection control programs and adhere to stringent regulatory guidelines, the demand for effective ultrasound probe disinfection solutions increases to minimize the risk of cross-contamination and maintain a safe healthcare environment.

Regulatory agencies and accreditation bodies impose strict standards and guidelines for the disinfection and sterilization of medical devices, including ultrasound probes. Compliance with regulatory requirements is essential for healthcare facilities to maintain accreditation status and demonstrate adherence to best practices in infection control. The need to comply with regulatory standards drives the demand for reliable and validated ultrasound probe disinfection solutions that meet regulatory requirements and ensure patient safety. Healthcare facilities prioritize patient safety and the delivery of high-quality care to enhance patient outcomes and satisfaction. Effective disinfection of ultrasound probes is integral to infection prevention strategies aimed at minimizing the risk of HAIs, reducing healthcare costs associated with treatment of infections, and improving overall patient experience. As healthcare providers prioritize patient safety and quality of care, the demand for ultrasound probe disinfection solutions that offer reliable and efficient disinfection becomes increasingly prominent. This factor will accelerate the demand of the Global Ultrasound Probe Disinfection Market.

Key Market Challenges

Concerns Regarding Probe Compatibility and Material Compatibility

Ultrasound probes are manufactured using a variety of materials, including plastics, metals, elastomers, and transducer components. Each material may have different properties and susceptibilities to damage from disinfectants or disinfection methods. Ensuring compatibility with diverse probe materials is essential to prevent damage that could compromise probe functionality or structural integrity. Incompatible disinfection solutions or methods may lead to degradation, discoloration, corrosion, or other forms of damage to ultrasound probes over time. Damage to probes

components can affect imaging quality, transducer performance, and overall reliability, potentially leading to costly repairs or premature replacement of probes. Healthcare facilities must select disinfection solutions and methods that are safe for use with the specific materials and components of their ultrasound probes to minimize the risk of damage. In addition to material compatibility, the effectiveness of disinfection solutions and methods is paramount in eliminating pathogens and ensuring patient safety.

Certain disinfection agents or techniques may be more effective against specific types of pathogens or resistant microorganisms than others. Healthcare facilities must carefully evaluate the efficacy of disinfection methods in achieving high-level disinfection (HLD) while preserving probe integrity and minimizing the risk of infection transmission. The lack of standardized disinfection protocols for ultrasound probes poses challenges for healthcare facilities in establishing consistent and reliable disinfection practices. Variability in disinfection protocols across different clinical settings or healthcare facilities may result in inconsistent disinfection outcomes and increased risk of infection transmission. Standardizing disinfection protocols based on evidence-based guidelines and best practices can help mitigate concerns regarding probe compatibility and material compatibility while ensuring effective disinfection and patient safety.

Cost Constraints and Budgetary Pressures

Implementing an effective ultrasound probe disinfection system often requires a significant initial investment. Healthcare facilities may face budgetary constraints when procuring disinfection equipment, such as automated disinfection machines, UV light systems, or hydrogen peroxide vapor systems, which can be costly to purchase and install. The high upfront costs of acquiring disinfection technology may pose challenges for healthcare facilities with limited financial resources. In addition to initial investment costs, healthcare facilities must consider ongoing maintenance expenses associated with ultrasound probe disinfection systems. Maintenance costs may include routine servicing, calibration, replacement of consumables, and repairs to ensure the continued functionality and reliability of disinfection equipment. Budgetary constraints may limit the ability of healthcare facilities to allocate funds for ongoing maintenance and upkeep of disinfection systems, potentially impacting system performance and disinfection outcomes.

Certain disinfection methods require the use of consumables, such as disinfectant solutions, wipes, cartridges, or filters, which contribute to recurring operational

expenses. Healthcare facilities must budget for the purchase of consumables to maintain uninterrupted supply and ensure proper disinfection practices. Cost considerations may influence the selection of disinfection methods or products based on their affordability and long-term cost-effectiveness. Healthcare facilities must prioritize resource allocation to address competing demands for limited financial resources across various departments and operational areas. Budgetary pressures may lead to trade-offs between investing in essential healthcare services, medical equipment, staffing, infrastructure improvements, and infection control measures, including ultrasound probe disinfection. Limited resources may hinder healthcare facilities' ability to invest in advanced disinfection technologies or comprehensive infection control programs, potentially compromising patient safety and quality of care.

Key Market Trends

Increasing Adoption of High-Level Disinfection (HLD) Systems

Healthcare facilities are increasingly prioritizing infection prevention and control measures to mitigate the risk of healthcare-associated infections (HAIs) and ensure patient safety. HLD systems offer a higher level of disinfection efficacy compared to low-level disinfection methods, making them essential for ensuring the thorough elimination of pathogens from ultrasound probes and reducing the risk of infection transmission. Regulatory agencies and healthcare accreditation bodies impose stringent guidelines and standards for the disinfection and sterilization of medical devices, including ultrasound probes. Compliance with regulatory requirements necessitates the implementation of HLD protocols for semi-critical and critical devices. Healthcare facilities adopt HLD systems to ensure adherence to regulatory standards, maintain accreditation status, and demonstrate a commitment to patient safety and quality of care.

HLD systems offer efficient and convenient disinfection solutions for ultrasound probes, minimizing downtime between procedures and optimizing workflow efficiency within healthcare settings. Automated HLD systems streamline disinfection processes, reduce reliance on manual labor, and minimize the risk of human error, enabling healthcare facilities to achieve consistent and reliable disinfection outcomes while maximizing resource utilization. HLD systems utilize advanced disinfection technologies, such as hydrogen peroxide vapor, ozone-based disinfection, ultraviolet (UV) light, or combination methods, to achieve a high level of pathogen eradication and ensure comprehensive disinfection of ultrasound probes. These technologies effectively target and destroy a broad spectrum of microorganisms, including bacteria, viruses, fungi, and

spores, thereby reducing the risk of infection transmission, and promoting patient safety.

Segmental Insights

Level of Disinfection Insights

The high-level disinfection segment is projected to experience rapid growth in the Global Ultrasound Probe Disinfection Market during the forecast period. Healthcare-associated infections pose a significant threat to patient safety and can result in substantial healthcare costs and adverse clinical outcomes. Ultrasound probes, which meet patients during invasive procedures or examinations, are potential vectors for the transmission of pathogens. As awareness of the role of medical devices in HAIs grows, healthcare facilities prioritize the implementation of rigorous disinfection protocols, including HLD, to mitigate the risk of infection transmission and safeguard patient health. The rise of multidrug-resistant organisms (MDROs) presents a formidable challenge to infection control efforts within healthcare settings. MDROs, such as methicillin-resistant *Staphylococcus aureus* (MRSA) and vancomycin-resistant *Enterococcus* (VRE), exhibit resistance to multiple antimicrobial agents and pose a heightened risk of healthcare-associated infections.

HLD methods capable of effectively eliminating MDROs and other resilient pathogens are essential for controlling the spread of infections and preventing outbreaks in healthcare facilities. The development of innovative disinfection technologies has expanded the capabilities of HLD systems, enabling rapid, efficient, and environmentally sustainable disinfection of ultrasound probes. Advanced HLD methods, such as hydrogen peroxide vapor, ozone-based disinfection, and ultraviolet (UV) light, offer enhanced efficacy against a wide range of pathogens while minimizing the use of harmful chemicals and reducing environmental impact. Healthcare facilities increasingly turn to these advanced HLD solutions to streamline disinfection workflows, improve efficiency, and ensure compliance with regulatory requirements.

Type of Probe Insights

The critical devices segment is projected to experience rapid growth in the Global Ultrasound Probe Disinfection Market during the forecast period. Critical devices, including ultrasound probes used in invasive procedures or those that encounter sterile body cavities, are subject to stringent regulatory standards and guidelines for disinfection and sterilization. Regulatory bodies mandate high-level disinfection (HLD) for critical devices to minimize the risk of healthcare-associated infections (HAIs) and

ensure patient safety. The increasing emphasis on compliance with regulatory standards drives the demand for advanced ultrasound probe disinfection solutions tailored to critical devices. Critical ultrasound probes pose a higher risk of cross-contamination due to their direct contact with sterile body cavities or invasive procedures. Inadequate disinfection of critical devices can lead to the transmission of pathogens and increase the risk of healthcare-associated infections among patients.

Healthcare facilities prioritize the implementation of rigorous disinfection protocols and invest in effective HLD systems to mitigate the risk of cross-contamination and enhance patient safety. The prevalence of healthcare-associated infections, including those caused by multidrug-resistant organisms, continues to rise globally. Critical devices, such as ultrasound probes used in invasive procedures, are implicated in the transmission of pathogens and the spread of HAIs within healthcare settings. As healthcare providers strive to minimize the incidence of HAIs and improve infection control practices, there is a growing demand for high-performance disinfection solutions capable of effectively eliminating a broad spectrum of pathogens from critical ultrasound probes. Ensuring patient safety and maintaining the quality of care are paramount priorities for healthcare providers. The effective disinfection of critical devices, including ultrasound probes, is essential to prevent healthcare-associated infections, reduce the risk of procedural complications, and safeguard patient outcomes. The adoption of advanced disinfection solutions for critical devices reflects healthcare facilities' commitment to upholding the highest standards of patient safety and quality of care.

Regional Insights

North America emerged as the dominant region in the Global Ultrasound Probe Disinfection Market in 2023. North America boasts advanced healthcare infrastructure, including well-established healthcare facilities, sophisticated medical technologies, and robust regulatory frameworks. The region's healthcare infrastructure facilitates the adoption of advanced ultrasound probe disinfection solutions and compliance with stringent regulatory standards for infection control. Regulatory authorities in North America, such as the Food and Drug Administration (FDA) in the United States and Health Canada, enforce stringent guidelines and standards for the disinfection of medical devices, including ultrasound probes.

Compliance with these regulations drives the adoption of high-quality disinfection technologies and ensures patient safety, contributing to the dominance of North America in the global market. North America has one of the highest healthcare expenditures globally, with significant investments in healthcare infrastructure, medical

equipment, and infection control measures. The region's substantial healthcare spending enables healthcare facilities to procure advanced ultrasound probe disinfection systems and invest in comprehensive infection prevention and control programs, thereby driving market growth.

Key Market Players

Virox technologies, Inc.

Sch?lke & Mayr GmbH

Parker Laboratories, Inc.

Metrex Research LLC.

BODE Chemie GmbH

Medevice Healthcare PVT LTD.

Borer Chemie AG

Tristel Solutions Ltd

Siemens Healthcare GmbH

FUJIFILM Healthcare Europe Holding AG

Report Scope:

In this report, the Global Ultrasound Probe Disinfection Market has been segmented into the following categories, in addition to the industry trends which have also been detailed below:

Ultrasound Probe Disinfection Market, By Products & Services:

Instruments

Consumables

Services

Ultrasound Probe Disinfection Market, By Level of Disinfection:

Low Level Disinfection

Intermediate Level Disinfection

High-Level Disinfection

Ultrasound Probe Disinfection Market, By Type of Probe:

Critical Devices

Noncritical Devices

Semicritical Devices

Ultrasound Probe Disinfection Market, By End User:

Hospitals & Diagnostic Imaging Centers

Ambulatory Care Centers

Academic & Research Institutes

Others

Ultrasound Probe Disinfection Market, By Region:

North America

United States

Canada

Mexico

Europe

Germany

United Kingdom

France

Italy

Spain

Asia-Pacific

China

Japan

India

Australia

South Korea

South America

Brazil

Argentina

Colombia

Middle East & Africa

South Africa

Saudi Arabia

UAE

Competitive Landscape

Company Profiles: Detailed analysis of the major companies present in the Global Ultrasound Probe Disinfection Market.

Available Customizations:

Global Ultrasound Probe Disinfection market report with the given market data, Tech Sci Research offers customizations according to a company's specific needs. The following customization options are available for the report:

Company Information

Detailed analysis and profiling of additional market players (up to five).

Contents

1. PRODUCT OVERVIEW

- 1.1. Market Definition
- 1.2. Scope of the Market
 - 1.2.1. Markets Covered
 - 1.2.2. Years Considered for Study
 - 1.2.3. Key Market Segmentations

2. RESEARCH METHODOLOGY

- 2.1. Objective of the Study
- 2.2. Baseline Methodology
- 2.3. Key Industry Partners
- 2.4. Major Association and Secondary Sources
- 2.5. Forecasting Methodology
- 2.6. Data Triangulation & Validation
- 2.7. Assumptions and Limitations

3. EXECUTIVE SUMMARY

- 3.1. Overview of the Market
- 3.2. Overview of Key Market Segmentations
- 3.3. Overview of Key Market Players
- 3.4. Overview of Key Regions/Countries
- 3.5. Overview of Market Drivers, Challenges, Trends

4. VOICE OF CUSTOMER

5. GLOBAL ULTRASOUND PROBE DISINFECTION MARKET OUTLOOK

- 5.1. Market Size & Forecast
 - 5.1.1. By Value
- 5.2. Market Share & Forecast
 - 5.2.1. By Products & Services (Instruments, Consumables, and Services)
 - 5.2.2. By Level of Disinfection (Low Level Disinfection, Intermediate Level Disinfection, and High-Level Disinfection)

5.2.3. By Type of Probe (Critical Devices, Noncritical Devices, and Semicritical Devices)

5.2.4. By End User (Hospitals & Diagnostic Imaging Centers, Ambulatory Care Centers, Academic & Research Institutes, and Others)

5.2.5. By Region

5.2.6. By Company (2023)

5.3. Market Map

6. NORTH AMERICA ULTRASOUND PROBE DISINFECTION MARKET OUTLOOK

6.1. Market Size & Forecast

6.1.1. By Value

6.2. Market Share & Forecast

6.2.1. By Products & Services

6.2.2. By Level of Disinfection

6.2.3. By Type of Probe

6.2.4. By End User

6.2.5. By Country

6.3. North America: Country Analysis

6.3.1. United States Ultrasound Probe Disinfection Market Outlook

6.3.1.1. Market Size & Forecast

6.3.1.1.1. By Value

6.3.1.2. Market Share & Forecast

6.3.1.2.1. By Products & Services

6.3.1.2.2. By Level of Disinfection

6.3.1.2.3. By Type of Probe

6.3.1.2.4. By End User

6.3.2. Canada Ultrasound Probe Disinfection Market Outlook

6.3.2.1. Market Size & Forecast

6.3.2.1.1. By Value

6.3.2.2. Market Share & Forecast

6.3.2.2.1. By Products & Services

6.3.2.2.2. By Level of Disinfection

6.3.2.2.3. By Type of Probe

6.3.2.2.4. By End User

6.3.3. Mexico Ultrasound Probe Disinfection Market Outlook

6.3.3.1. Market Size & Forecast

6.3.3.1.1. By Value

6.3.3.2. Market Share & Forecast

- 6.3.3.2.1. By Products & Services
- 6.3.3.2.2. By Level of Disinfection
- 6.3.3.2.3. By Type of Probe
- 6.3.3.2.4. By End User

7. EUROPE ULTRASOUND PROBE DISINFECTION MARKET OUTLOOK

7.1. Market Size & Forecast

7.1.1. By Value

7.2. Market Share & Forecast

7.2.1. By Products & Services

7.2.2. By Level of Disinfection

7.2.3. By Type of Probe

7.2.4. By End User

7.2.5. By Country

7.3. Europe: Country Analysis

7.3.1. Germany Ultrasound Probe Disinfection Market Outlook

7.3.1.1. Market Size & Forecast

7.3.1.1.1. By Value

7.3.1.2. Market Share & Forecast

7.3.1.2.1. By Products & Services

7.3.1.2.2. By Level of Disinfection

7.3.1.2.3. By Type of Probe

7.3.1.2.4. By End User

7.3.2. United Kingdom Ultrasound Probe Disinfection Market Outlook

7.3.2.1. Market Size & Forecast

7.3.2.1.1. By Value

7.3.2.2. Market Share & Forecast

7.3.2.2.1. By Products & Services

7.3.2.2.2. By Level of Disinfection

7.3.2.2.3. By Type of Probe

7.3.2.2.4. By End User

7.3.3. Italy Ultrasound Probe Disinfection Market Outlook

7.3.3.1. Market Size & Forecast

7.3.3.1.1. By Value

7.3.3.2. Market Share & Forecast

7.3.3.2.1. By Products & Services

7.3.3.2.2. By Level of Disinfection

7.3.3.2.3. By Type of Probe

- 7.3.3.2.4. By End User
- 7.3.4. France Ultrasound Probe Disinfection Market Outlook
 - 7.3.4.1. Market Size & Forecast
 - 7.3.4.1.1. By Value
 - 7.3.4.2. Market Share & Forecast
 - 7.3.4.2.1. By Products & Services
 - 7.3.4.2.2. By Level of Disinfection
 - 7.3.4.2.3. By Type of Probe
 - 7.3.4.2.4. By End User
- 7.3.5. Spain Ultrasound Probe Disinfection Market Outlook
 - 7.3.5.1. Market Size & Forecast
 - 7.3.5.1.1. By Value
 - 7.3.5.2. Market Share & Forecast
 - 7.3.5.2.1. By Products & Services
 - 7.3.5.2.2. By Level of Disinfection
 - 7.3.5.2.3. By Type of Probe
 - 7.3.5.2.4. By End User

8. ASIA-PACIFIC ULTRASOUND PROBE DISINFECTION MARKET OUTLOOK

- 8.1. Market Size & Forecast
 - 8.1.1. By Value
- 8.2. Market Share & Forecast
 - 8.2.1. By Products & Services
 - 8.2.2. By Level of Disinfection
 - 8.2.3. By Type of Probe
 - 8.2.4. By End User
 - 8.2.5. By Country
- 8.3. Asia-Pacific: Country Analysis
 - 8.3.1. China Ultrasound Probe Disinfection Market Outlook
 - 8.3.1.1. Market Size & Forecast
 - 8.3.1.1.1. By Value
 - 8.3.1.2. Market Share & Forecast
 - 8.3.1.2.1. By Products & Services
 - 8.3.1.2.2. By Level of Disinfection
 - 8.3.1.2.3. By Type of Probe
 - 8.3.1.2.4. By End User
 - 8.3.2. India Ultrasound Probe Disinfection Market Outlook
 - 8.3.2.1. Market Size & Forecast

- 8.3.2.1.1. By Value
- 8.3.2.2. Market Share & Forecast
 - 8.3.2.2.1. By Products & Services
 - 8.3.2.2.2. By Level of Disinfection
 - 8.3.2.2.3. By Type of Probe
 - 8.3.2.2.4. By End User
- 8.3.3. Japan Ultrasound Probe Disinfection Market Outlook
 - 8.3.3.1. Market Size & Forecast
 - 8.3.3.1.1. By Value
 - 8.3.3.2. Market Share & Forecast
 - 8.3.3.2.1. By Products & Services
 - 8.3.3.2.2. By Level of Disinfection
 - 8.3.3.2.3. By Type of Probe
 - 8.3.3.2.4. By End User
- 8.3.4. South Korea Ultrasound Probe Disinfection Market Outlook
 - 8.3.4.1. Market Size & Forecast
 - 8.3.4.1.1. By Value
 - 8.3.4.2. Market Share & Forecast
 - 8.3.4.2.1. By Products & Services
 - 8.3.4.2.2. By Level of Disinfection
 - 8.3.4.2.3. By Type of Probe
 - 8.3.4.2.4. By End User
- 8.3.5. Australia Ultrasound Probe Disinfection Market Outlook
 - 8.3.5.1. Market Size & Forecast
 - 8.3.5.1.1. By Value
 - 8.3.5.2. Market Share & Forecast
 - 8.3.5.2.1. By Products & Services
 - 8.3.5.2.2. By Level of Disinfection
 - 8.3.5.2.3. By Type of Probe
 - 8.3.5.2.4. By End User

9. SOUTH AMERICA ULTRASOUND PROBE DISINFECTION MARKET OUTLOOK

- 9.1. Market Size & Forecast
 - 9.1.1. By Value
- 9.2. Market Share & Forecast
 - 9.2.1. By Products & Services
 - 9.2.2. By Level of Disinfection
 - 9.2.3. By Type of Probe

- 9.2.4. By End User
- 9.2.5. By Country
- 9.3. South America: Country Analysis
 - 9.3.1. Brazil Ultrasound Probe Disinfection Market Outlook
 - 9.3.1.1. Market Size & Forecast
 - 9.3.1.1.1. By Value
 - 9.3.1.2. Market Share & Forecast
 - 9.3.1.2.1. By Products & Services
 - 9.3.1.2.2. By Level of Disinfection
 - 9.3.1.2.3. By Type of Probe
 - 9.3.1.2.4. By End User
 - 9.3.2. Argentina Ultrasound Probe Disinfection Market Outlook
 - 9.3.2.1. Market Size & Forecast
 - 9.3.2.1.1. By Value
 - 9.3.2.2. Market Share & Forecast
 - 9.3.2.2.1. By Products & Services
 - 9.3.2.2.2. By Level of Disinfection
 - 9.3.2.2.3. By Type of Probe
 - 9.3.2.2.4. By End User
 - 9.3.3. Colombia Ultrasound Probe Disinfection Market Outlook
 - 9.3.3.1. Market Size & Forecast
 - 9.3.3.1.1. By Value
 - 9.3.3.2. Market Share & Forecast
 - 9.3.3.2.1. By Products & Services
 - 9.3.3.2.2. By Level of Disinfection
 - 9.3.3.2.3. By Type of Probe
 - 9.3.3.2.4. By End User

10. MIDDLE EAST AND AFRICA ULTRASOUND PROBE DISINFECTION MARKET OUTLOOK

- 10.1. Market Size & Forecast
 - 10.1.1. By Value
- 10.2. Market Share & Forecast
 - 10.2.1. By Products & Services
 - 10.2.2. By Level of Disinfection
 - 10.2.3. By Type of Probe
 - 10.2.4. By End User
 - 10.2.5. By Country

10.3. MEA: Country Analysis

10.3.1. South Africa Ultrasound Probe Disinfection Market Outlook

10.3.1.1. Market Size & Forecast

10.3.1.1.1. By Value

10.3.1.2. Market Share & Forecast

10.3.1.2.1. By Products & Services

10.3.1.2.2. By Level of Disinfection

10.3.1.2.3. By Type of Probe

10.3.1.2.4. By End User

10.3.2. Saudi Arabia Ultrasound Probe Disinfection Market Outlook

10.3.2.1. Market Size & Forecast

10.3.2.1.1. By Value

10.3.2.2. Market Share & Forecast

10.3.2.2.1. By Products & Services

10.3.2.2.2. By Level of Disinfection

10.3.2.2.3. By Type of Probe

10.3.2.2.4. By End User

10.3.3. UAE Ultrasound Probe Disinfection Market Outlook

10.3.3.1. Market Size & Forecast

10.3.3.1.1. By Value

10.3.3.2. Market Share & Forecast

10.3.3.2.1. By Products & Services

10.3.3.2.2. By Level of Disinfection

10.3.3.2.3. By Type of Probe

10.3.3.2.4. By End User

11. MARKET DYNAMICS

11.1. Drivers

11.2. Challenges

12. MARKET TRENDS & DEVELOPMENTS

12.1. Merger & Acquisition (If Any)

12.2. Product Launches (If Any)

12.3. Recent Developments

13. PORTER'S FIVE FORCES ANALYSIS

- 13.1. Competition in the Industry
- 13.2. Potential of New Entrants
- 13.3. Power of Suppliers
- 13.4. Power of Customers
- 13.5. Threat of Substitute Product

14. COMPETITIVE LANDSCAPE

- 14.1. Virox technologies, Inc.
 - 14.1.1. Business Overview
 - 14.1.2. Company Snapshot
 - 14.1.3. Products & Services
 - 14.1.4. Financials (As Reported)
 - 14.1.5. Recent Developments
 - 14.1.6. Key Personnel Details
 - 14.1.7. SWOT Analysis
- 14.2. Sch?lke & Mayr GmbH
- 14.3. Parker Laboratories, Inc.
- 14.4. Metrex Research LLC.
- 14.5. BODE Chemie GmbH
- 14.6. Medevice Healthcare PVT LTD.
- 14.7. Borer Chemie AG
- 14.8. Tristel Solutions Ltd
- 14.9. Siemens Healthcare GmbH
- 14.10.FUJIFILM Healthcare Europe Holding AG

15. STRATEGIC RECOMMENDATIONS

16. ABOUT US & DISCLAIMER

I would like to order

Product name: Ultrasound Probe Disinfection Market - Global Industry Size, Share, Trends, Opportunity, and Forecast, Segmented By Products & Services (Instruments, Consumables, and Services), By Level of Disinfection (Low Level Disinfection, Intermediate Level Disinfection, and High Level Disinfection), By Type of Probe (Critical Devices, Noncritical Devices, and Semicritical Devices), By End User (Hospitals & Diagnostic Imaging Centers, Ambulatory Care Centers, Academic & Research Institutes, and Others), By Region, and By Competition, 2019-2029F

Product link: <https://marketpublishers.com/r/U68E95693C18EN.html>

Price: US\$ 4,900.00 (Single User License / Electronic Delivery)

If you want to order Corporate License or Hard Copy, please, contact our Customer Service:

info@marketpublishers.com

Payment

To pay by Credit Card (Visa, MasterCard, American Express, PayPal), please, click button on product page <https://marketpublishers.com/r/U68E95693C18EN.html>

To pay by Wire Transfer, please, fill in your contact details in the form below:

First name:
Last name:
Email:
Company:
Address:
City:
Zip code:
Country:
Tel:
Fax:
Your message:

****All fields are required**

Customer signature _____

Please, note that by ordering from marketpublishers.com you are agreeing to our Terms & Conditions at <https://marketpublishers.com/docs/terms.html>

To place an order via fax simply print this form, fill in the information below and fax the completed form to +44 20 7900 3970